

Accreditation Certificate

Reagecon Diagnostics Ltd.

Shannon Free Zone, Shannon, Co. Clare

Testing Laboratory

Registration number: 264T


is accredited by the Irish National Accreditation Board (INAB) to undertake testing as detailed in the Schedule bearing the Registration Number detailed above, in compliance with the International Standard **ISO/IEC 17025:2005 2nd Edition** “*General Requirements for the Competence of Testing and Calibration Laboratories*”
(This Certificate must be read in conjunction with the Annexed Schedule of Accreditation)


Date of award of accreditation: 12:07:2010

Date of last renewal of accreditation: 14:07:2016

Expiry date of this certificate of accreditation: 14:07:2021

This Accreditation shall remain in force until further notice subject to continuing compliance with INAB accreditation criteria, ISO/IEC 17025 and any further requirements specified by the Irish National Accreditation Board.


Manager: _____
Dr Adrienne Duff


Chairperson: _____
Mr Tom O'Neill

Issued on 14 July 2016

Organisations are subject to annual surveillance and are re-assessed every five years. The renewal date on this Certificate confirms the latest date of renewal of accreditation. To confirm the validity of this Certificate, please contact the Irish National Accreditation Board.
The INAB is a signatory of the European co-operation for Accreditation (EA) Testing Multilateral Agreement (MLA) and the International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Arrangement.

Schedule of Accreditation



(Annex to Accreditation Certificate)

Permanent Laboratory:

Category A

REAGECON DIAGNOSTICS LTD

Chemical Testing Laboratory

Initial Registration Date : 12-July-2010
Postal Address: Shannon Free Zone
(Address of other locations as they apply) Shannon
Co Clare
Telephone: +353 (61) 472622
Fax: +353 (61) 472642
E-mail: John.barron@reagecon.ie
Contact Name: John Barron
Facilities: Normally not available for Public testing

Schedule of Accreditation



Permanent Laboratory:
Category A

THE IRISH NATIONAL ACCREDITATION BOARD (INAB) is the Irish body for the accreditation of organisations including laboratories.

Laboratory accreditation is available to testing and calibration facilities operated by manufacturing organisations, government departments, educational institutions and commercial testing/calibration services. Indeed, any organisation involved in testing, measurement or calibration in any area of technology can seek accreditation for the work it is undertaking.

Each accredited laboratory has been assessed by skilled specialist assessors and found to meet criteria which are in compliance with ISO/IEC 17025 or ISO 15189 (medical laboratories). Frequent audits, together with periodic inter-laboratory test programmes, ensure that these standards of operation are maintained.

Testing and Calibration Categories:

- Category A:** Permanent laboratory calibration and testing where the laboratory is erected on a fixed location for a period expected to be greater than three years.
- Category B:** Site calibration and testing that is performed by staff sent out on site by a permanent laboratory that is accredited by the Irish National Accreditation Board.
- Category C:** Site calibration and testing that is performed in a site/mobile laboratory or by staff sent out by such a laboratory, the operation of which is the responsibility of a permanent laboratory accredited by the Irish National Accreditation Board.
- Category D:** Site calibration and testing that is performed on site by individuals and organisations that do not have a permanent calibration/testing laboratory. Testing may be performed using
- (a) portable test equipment
 - (b) a site laboratory
 - (c) a mobile laboratory or
 - (d) equipment from a mobile or site laboratory

Standard Specification or Test Procedure Used:

The standard specification or test procedure that is accredited is the issue that is current on the date of the most recent visit, unless otherwise stated.

Glossary of Terms

Facilities:

- Public calibration/testing service:** Commercial operations which actively seek work from others.
- Conditionally available for public calibration/testing:** Established for another primary purpose but, more commonly than not, is available for outside work.
- Normally not available for public calibration/testing:** Unavailable for public calibration/testing more often than not.

Laboratory users wishing to obtain assurance that calibration or test results are reliable and carried out to the Irish National Accreditation Board criteria should insist on receiving an accredited calibration certificate or test report. Users should contact the laboratory directly to ensure that this scope of accreditation is current. INAB will, on request, verify the status and scope.

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 .11 Laboratory Reagents Chemical Tests Standard Buffer Solutions Solid Buffers Aqueous General Reagents	 pH 0.98 to 13.05 pH 3.98 to 10.02 pH 0.98 to 13.05	Documented In-House method by Electrometry TPPHB TPPHC TPPHG
795 .11 Laboratory Reagents Chemical Tests Sodium Thiosulphate by redox reaction Iodine by redox reaction Acetous Perchloric Acid Silver Nitrate by argentometric titration EDTA by compleximetric titration	 0.0099M to 1.001M 0.0499M to 0.5005M 0.0998M to 0.1002M 0.0499M to 1.002M 0.00998M to 1.002M	Documented In-House Methods using titrimetry, based on ASTM E200-08 (See Note 1) TPATRX1 TPATRX2 TPAHCLO TPAGNO TPEDTA

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 Laboratory Reagents .11 Chemical Tests Chloride Solutions	Chloride Solutions 0.0099M to 0.0905M Chloride Solutions 0.0998M to 4.008M	Documented In-House volumetric methods based on methods from Vogel: Quantitative Inorganic Analysis, 4th Edition:- TPATPPT1
Acid Solutions <i>Monobasic Acids</i> <i>Dibasic Acids</i>	Monobasic Acids 0.0249M - 10.01M 0.0099M - 0.0240M Dibasic Acids 0.0249M - 10.01M	TPATA Acid Base Titration
Base Solutions <i>Hydroxide Ion</i>	Hydroxide Ion 0.0199M - 0.0491M 0.0499M - 10.01M	TPATB Acid Base Titration
General Reagents	Conductivity 4.95 to 505,000 $\mu\text{S}/\text{cm}$ @ 25 deg C Conductivity at 1.25 to 1.35 $\mu\text{S}/\text{cm}$ @ 25 deg C	TPCOND

Note 1. ASTM: American Society for Testing and Materials

Scope of Accreditation



Reagecon Diagnostics Ltd

Permanent Laboratory:

Chemical Testing Laboratory

Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
<p>795 Laboratory Reagents .11 Chemical Tests</p>	<p>Analysis of elements in the range 49µg/Kg to 1.48mg/Kg for the following:- Beryllium Boron Sodium Magnesium Aluminium Phosphorus Potassium Calcium Vanadium Chromium Manganese Iron Cobalt Nickel Copper Zinc Gallium Arsenic Selenium Rubidium Strontium Silver Cadmium Caesium</p>	<p>Documented In-House method TPICP by ICP-MS, based on US EPA Method 6020A</p>

Scope of Accreditation



Reagecon Diagnostics Ltd

Permanent Laboratory:

Chemical Testing Laboratory

Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 Laboratory Reagents .11 Chemical Tests	Analysis of elements in the range 49µg/Kg to 1.48mg/Kg for the following:- Barium Lanthanum Cerium Praseodymium Neodymium Samarium Europium Gadolinium Dysprosium Holmium Erbium Thulium Ytterbium Thallium Lead Thorium Uranium	Documented In-House method TPICP by ICP-MS, based on US EPA Method 6020A

Scope of Accreditation



Reagecon Diagnostics Ltd

Permanent Laboratory:

Chemical Testing Laboratory

Category A

INAB Classification number (P9) Materials/products tested		Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 .11	Laboratory Reagents Chemical Tests	Analysis of elements in the range 49µg/Kg to 1.48mg/Kg for the following:- Hofnium Molybdenum Niobium Antimony Tin Tantalum Tellurium Titanium Tungsten Zirconium Silicon range 86µg/Kg to 1.48mg/Kg	Documented In-House method TPICP by ICP-MS, based on US EPA Method 6020A
795 .11	Laboratory Reagents Chemical Tests	Analysis of elements metals for the following:- Sulphur:- range 1.50mg/Kg to 3.50mg/Kg Metals over the range 47µg/Kg to 1.41mg/Kg Ruthenium Palladium Rhenium Osmium Iridium Platinum Gold	Documented In-House method TPCIPPR by ICP-MS, based on US EPA Method 6020A

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested		Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 .11	Laboratory Reagents Chemical Tests	Analysis of Mercury over the range 47µg/Kg to 1.41mg/Kg	Documented In-House method TPICPHG by ICP-MS, based on US EPA Method 6020A
795 .11	Laboratory Reagents Chemical Tests	Analysis of elements for the following elements:- Lithium (Li 6) range 48µg/Kg to 1.43mg/Kg Lithium (Li 7) range 50µg/Kg to 1.49mg/Kg Scandium - range 48µg/Kg to 1.43mg/Kg Germanium - range 48µg/Kg to 1.43mg/Kg Yttrium - range 50µg/Kg to 1.49mg/Kg Rhodium - range 48µg/Kg to 1.43mg/Kg Indium - range 48µg/Kg to 1.43mg/Kg Terbium - range 48µg/Kg to 1.43mg/Kg Lutetium - range 48µg/Kg to 1.43mg/Kg Bismuth - range 48µg/Kg to 1.43mg/Kg	Documented In-House method TPCIS by ICP-MS, based on US EPA Method 6020A

Scope of Accreditation



Reagecon Diagnostics Ltd

Permanent Laboratory:

Category A

Chemical Testing Laboratory

INAB Classification number (P9) Materials/products tested		Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795 .11	Laboratory Reagents Chemical Tests	Determination of Total Acid Number from 0.09 mg/g KOH to 4.70 mg/g KOH	Documented in-House test method TP-TAN in accordance with ASTM D664, Test Method A
795 .11	Laboratory Reagents Chemical Tests	Determination of Total Base Number from 0.9 mg/g KOH to 75 mg/g KOH	Documented in-House test method TPTBN in accordance with ASTM D2896, Test Method A
795 .11	Laboratory Reagents Chemical Tests	Determination of the following metals in the concentration range 100-1000 mg/L for the following elements: Bismuth Calcium Cadmium Cobalt Copper Indium Iron Manganese Nickel Scandium Tin Vanadium Lead	Documented in-House test method TPCOMPLEX by Compleximetric (EDTA) titration based on the primary methods described in Applied Complexometry- Pribil, Rudolf - Pergamon Press - 1982

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.12 Laboratory Reagents Physico-Chemical Test	Density of liquid materials range 0.65 to 1.034 g/ml	Documented In-House method TPDMA5000M by vibrational methods, based on ASTM D4052
795.12 Physico-Chemical Test	Refractive Index range 1.33310 to 1.65812nD Brix Value range 5% to 60% wt/wt	Documented In-House method TPRIA 01, based on OIML R142
795.12 Physico-Chemical Test	Osmolality range 50 to 3000 mOsm/kg H ₂ O	Documented In-House method TPOSM-1500 and TPOSM-3000, based on USP - 785 and EP - 7.0
795.11 Laboratory Reagents Chemical Tests	Determination of (TOC) Total Organic Carbon Range 500 µg/l to 50.0mg/l C	Documented In-House Method TPTIC
795.11 Laboratory Reagents Chemical Tests	Determination of (TIC) Total Inorganic Carbon Range 500 µg/l to 50.0 mg/l C	Documented In-House Method TPTOC
795 .11 Laboratory Reagents Chemical Tests	Determination of Total Base Number (TBN) 0.9 - 75 mg/g KOH	Documented In-House Method based on ASTM D4739-11 TPTBNUO / Automatic Titration
795 .12 Laboratory Reagents Physical Tests	Determination of Density 0.63 - 1.63 g/ml in the working temperature 15°C to 50°C	ASTM D1480-15 TPPYC / Bingham Pycnometer

Scope of Accreditation



Reagecon Diagnostics Ltd

Chemical Testing Laboratory

Permanent Laboratory:

Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
<p>795 .12 Laboratory Reagents Physical Tests</p>	<p>Viscosity Measurement Kinematic Viscosity and Dynamic Viscosity</p> <p>Temperature Range +20°C to +60°C - Kinematic Viscosity Range 0.85 to 110000 cSt</p> <p>Temperature Range +20°C to +60°C - Dynamic Viscosity Range 0.55 to 99000 mPa.s</p>	<p>ASTM D2162 - 14 Master Ubbelohde Capillary Viscometers / Viscometry</p>

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.11 Laboratory Reagents Physical Tests	Analysis of Phenols by GC/MS in the range of 20 to 2000 µg/ml 4-Chloro-3-methyl phenol 2-Chlorophenol 2,4-Dichlorophenol 2,6-Dichlorophenol 2,4-Dimethyl phenol 4,6-Dinitro-2-methyl phenol 2,4-Dinitrophenol 2-(1- Methylpropyl)-4,6-dinitrophenol 2-Methyl phenol 3-Methyl phenol 2-Nitrophenol 4-Nitrophenol Pentachlorophenol 2,3,4,6-Tetrachlorophenol 2,4,5-Trichlorophenol 2,4,6-Trichlorophenol	In-House Method TPGCMS02 by GC/MS

Scope of Accreditation



Reagecon Diagnostics Ltd

Permanent Laboratory:

Chemical Testing Laboratory

Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.11 Laboratory Reagents Physical Tests	Analysis of Volatile Organic Carbons by GC/MS in the range of 20 to 2000 µg/ml Benzene Bromobenzene Bromochloromethane Bromodichloromethane Bromoform n-Butylbenzene sec-Butylbenzene tert-Butylbenzene Carbon Tetrachloride Chloroform Chlorobenzene 2-Chlorotoluene 4-Chlorotoluene 1,2-Dibromo-3-chloropropane Dibromoethane Dibromomethane Dibromochloromethane 1,2-Dichlorobenzene 1,3-Dichlorobenzene 1,4-Dichlorobenzene 1,1-Dichloroethene Cis-1,2-Dichloroethane Trans-1,2-dichloroethane 1,1-Dichloroethane 1,2-Dichloroethane	In-House Method TPGCMS01 by GC/MS

Scope of Accreditation



Reagecon Diagnostics Ltd

Chemical Testing Laboratory

Permanent Laboratory:

Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.11 Laboratory Reagents Physical Tests	Analysis of Volatile Organic Carbons by GC/MS in the range of 20 to 2000 µg/ml Dichloromethane 1,2-Dichloropropane 2,2-Dichloropropane 1,3-Dichloropropane 1,1-Dichloropropene cis-1,2-Dichloropropene trans-1,2-Dichloropropene Ethylbenzene Hexachlorobutadiene Isopropylbenzene 4-Isopropyltoluene Naphthalene n-Propylbenzene Styrene Trichloroethene 1,1,1-Trichloroethene Toluene 1,1,2-Trichloroethane Tetrachloroethane 1,1,1,2-Tetrachloroethane 1,2,3-Trichloropropane 1,2,4-Trimethylbenzene 1,3,5-Trimethylbenzene 1,2,3-Trichlorobenzene 1,2,4-Trichlorobenzene (m+p) Xylenes o-Xylene	In-House Method TPGCMS01 by GC/MS

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.11 Laboratory Reagents Physical Tests	Analysis of Pesticides by GC/MS in the range of 20 to 2000 µg/ml Aldrin Atrazine Gamma-Chlordane Chlorobenzilate Chloroneb Chlorothalonil DCPA p,p'-DDD p,p'-DDE p,p'-DDT Dieldrin Endosulfan 1 Endosulfan 2 Endosulfan sulfate Endrin Endrin aldehyde Etridiazole Alpha HCH Beta HCH Gamma-HCH (Lindane) Delta-HCH Heptachlor Heptachlor epoxide Hexachlorobenzene Methoxychlor cis Permethrin trans Permethrin Propachlor	In-House Method TPGCMS04 by GC/MS.

Scope of Accreditation



Reagecon Diagnostics Ltd Chemical Testing Laboratory

Permanent Laboratory:
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
795.11 Laboratory Reagents Physical Tests	Analysis of Pesticides by GC/MS in the range of 20 to 2000 µg/ml Simazine Trifluralin	In-House Method TPGCMS04 by GC/MS.
795.11 Laboratory Reagents Physical Tests	Analysis of PAH by GC/MS in the range of 10 to 2000 µg/ml Acenaphthene Acenaphthylene Anthracene Benzo(a)anthracene Benzo(a)pyrene Benzo(b)fluoranthene Benzo(g,h,i)perylene Benzo(k)fluoranthene Chrysene Dibenz(a,h)anthracene Fluorene Fluoranthene Indeno(1,2,3-cd)pyrene Naphthalene Phenanthrene Pyrene	In-House Method TPGCMS03 by GC/MS