

Accreditation Certificate

Complete Calibrations

Unit 33, Silverheights Avenue, Mayfield, Co Cork

Calibration Laboratory

Registration number: 282C

is accredited by the Irish National Accreditation Board (INAB) to undertake calibration as detailed in the Schedule bearing the Registration Number detailed above, in compliance with the International Standard ISO/IEC 17025:2005 2nd Edition

“General Requirements for the Competence of Testing and Calibration Laboratories”

(This Certificate must be read in conjunction with the annexed Schedule of Accreditation)

Date of award of accreditation: 15:02:2011


Date of last renewal of accreditation: 15:02:2011

Expiry date of this certificate of accreditation: 15:02:2016

This Accreditation shall remain in force until further notice subject to continuing compliance with INAB accreditation criteria, ISO/IEC 17025 and any further requirements specified by the Irish National Accreditation Board.

Manager: 

Dr Adrienne Duff

Chairperson: 

Mr Tom O'Neill

Issued on 15 February 2011

Organisations are subject to annual surveillance and are re-assessed every five years. The renewal date on this Certificate confirms the latest date of renewal of accreditation. To confirm the validity of this Certificate, please contact the Irish National Accreditation Board.

INAB is a signatory of the European co-operation for Accreditation (EA) Testing Multilateral Agreement (MLA) and the International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Arrangement.

Schedule of Accreditation



(Annex to Accreditation Certificate)

Permanent Laboratory:
Category A

COMPLETE CALIBRATIONS

Mass Calibration Laboratory

Initial Accreditation Date : 15-February-2011
Postal Address: Unit 33, Silverheights Avenue, Mayfield, Co Cork
Telephone: +353 (21) 4506572
Website: www.compcals.com
E-mail: compcals@gmail.com
Contact Name: Mr Darren O'Sullivan
Facilities: Public calibration service

Schedule of Accreditation



Permanent Laboratory:
 Category A

THE IRISH NATIONAL ACCREDITATION BOARD (INAB) is the Irish body for the accreditation of organisations including laboratories.

Laboratory accreditation is available to testing and calibration facilities operated by manufacturing organisations, government departments, educational institutions and commercial testing/calibration services. Indeed, any organisation involved in testing, measurement or calibration in any area of technology can seek accreditation for the work it is undertaking.

Each accredited laboratory has been assessed by skilled specialist assessors and found to meet criteria which are in compliance with ISO/IEC 17025 or ISO/IEC 15189 (medical laboratories). Frequent audits, together with periodic inter-laboratory test programmes, ensure that these standards of operation are maintained.

Calibration Categories:

- Category A:** Permanent calibration laboratory where the laboratory is erected on a fixed location for a period expected to be greater than three years.
- Category B:** Site calibration that is performed by staff sent out on site by a permanent laboratory that is accredited by the Irish National Accreditation Board.
- Category C:** Site calibration that is performed in a site/mobile laboratory or by staff sent out by such a laboratory, the operation of which is the responsibility of a permanent laboratory accredited by the Irish National Accreditation Board.
- Category D:** Site calibration that is performed on site by individuals and organisations that do not have a permanent calibration laboratory. Calibration may be performed using
- (a) portable test equipment
 - (b) a site laboratory
 - (c) a mobile laboratory or
 - (d) equipment from a mobile or site laboratory

Standard Specification or Calibration Procedure Used:

The standard specification or calibration procedure that is accredited is the issue that is current on the date of the most recent visit, unless otherwise stated.

Glossary of Terms

Facilities:

- Public calibration service:** Commercial operations which actively seek work from others.
- Conditionally available for public calibration:** Established for another primary purpose but, more commonly than not, is available for outside work.
- Normally not available for public calibration:** Unavailable for public calibration more often than not.

Laboratory users wishing to obtain assurance that calibration results are reliable and carried out to the Irish National Accreditation Board criteria should insist on receiving an accredited calibration certificate. Users should contact the laboratory directly to ensure that this schedule of accreditation is current. INAB will on request verify the status and scope.

Scope of Accreditation



Complete Calibrations

Permanent Laboratory:
Category A

Mass Calibration Laboratory

(Nominal temperature for calibration work: $20 \pm 1.0^{\circ}\text{C}$)

INAB Classification number (P9)	Measured quantity	Range of measurement	Calibration and measurement capability expressed as an uncertainty *	Method and remarks
120	Mass	Range-Nominal value in g		
.02	Industrial Mass Standards	(g)	(mg)	Based on documented in-house methods SOP 0015 section 3.2 in compliance with O.I.M.L. Class E2 standards from 1mg to 2000g inclusive
	F2	20,000	60	
	F1	10,000	10	
	F1	5,000	5	
	E2	2,000	1.0	
	E2	1,000	0.533	
	E2	500	0.267	
	E2	200	0.100	
	E2	100	0.053	
	E2	50	0.033	
	E2	20	0.027	
	E2	10	0.020	
	E2	5	0.017	
	E2	2	0.013	
	E2	1	0.010	

*** Notes:**

- Intermediate values can be calibrated with an uncertainty interpolated from the next higher and lower values in the table above.
- Calibration can be given in other units as required.
- In accordance with INAB policy, uncertainties are calculated for an estimated confidence level of not less than 95%.
- Calibration and measurement capability expressed as an uncertainty (\pm) to be reported in compliance with EA-4/02, "Expression of the Uncertainty of Measurement in Calibration".

Scope of Accreditation



Permanent Laboratory:
Category A

Complete Calibrations

Mass Calibration Laboratory

(Nominal temperature for calibration work: $20 \pm 1.0^{\circ}\text{C}$)

INAB Classification number (P9)	Measured quantity	Range of measurement	Calibration and measurement capability expressed as an uncertainty *	Method and remarks
120	Mass	Range-Nominal value in g		
.02	Industrial Mass Standards	(g)	(mg)	Based on documented in-house methods SOP 0015 section 3.2 in compliance with O.I.M.L. Class E2 standards from 1mg to 2000g inclusive
	E2	0.5	0.0083	
	E2	0.2	0.0067	
	E2	0.1	0.0053	
	E2	0.05	0.0040	
	E2	0.02	0.0033	
	E2	0.01	0.0027	Class F1 standards
	E2	0.005	0.0020	from 1mg to 10,000g inclusive
	E2	0.002	0.0020	
	E2	0.001	0.0020	Class F2, M1, M2, M3 standards from 1mg to 20,000g inclusive

*** Notes:**

- Intermediate values can be calibrated with an uncertainty interpolated from the next higher and lower values in the table above.
- Calibration can be given in other units as required.
- In accordance with INAB policy, uncertainties are calculated for an estimated confidence level of not less than 95%.
- Calibration and measurement capability expressed as an uncertainty (\pm) to be reported in compliance with EA-4/02, "Expression of the Uncertainty of Measurement in Calibration".