

# Accreditation Certificate

## Adelaide & Meath Hospital incorporating The National Children's Hospital

Blood transfusion

Testing Laboratory

Registration number: 213MT

is accredited by the Irish National Accreditation Board (INAB) to undertake testing as detailed in the Schedule bearing the Registration Number detailed above, in compliance with the International Standard ISO/IEC 15189:2007 2nd Edition

*"Medical laboratories - Particular requirements for quality and competence"*  
*(This Certificate must be read in conjunction with the Annexed Schedule of Accreditation)*

---

Date of award of accreditation: 17:02:2009

Date of last renewal of accreditation: 17:02:2009

Expiry date of this certificate of accreditation: 17:02:2014

---

This Accreditation shall remain in force until further notice subject to continuing compliance with INAB accreditation criteria, ISO/IEC 15189 and any further requirements specified by the Irish National Accreditation Board.

Manager:  Chairperson:   
Dr Adrienne Duff Dr Máire Walsh

Issued on 17 February 2009

Organisations are subject to annual surveillance and are re-assessed every five years. The renewal date on this Certificate confirms the latest date of renewal of accreditation. To confirm the validity of this Certificate, please contact the Irish National Accreditation Board.

The INAB is a signatory of the European co-operation for Accreditation (EA) Testing Multilateral Agreement (MLA) and the International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Arrangement.

# Schedule of Accreditation



(Annex to Accreditation Certificate)

Permanent Laboratory:  
Category A

## ADELAIDE & MEATH HOSPITAL (INC NCH) Dublin 24

### Blood Transfusion Testing Laboratory

*Initial Registration Date :* 17-February-2009  
*Postal Address:* Tallaght  
*(Address of other locations as they apply)* Dublin 24  
*Telephone:* +353 (1) 4143910  
*Fax:* +353 (1) 4143996  
*E-mail:* [gerry.judge@amnch.ie](mailto:gerry.judge@amnch.ie)  
*Contact Name:* Mr Gerry Judge  
*Facilities:* Public testing service

# Schedule of Accreditation



Permanent Laboratory:  
 Category A

THE IRISH NATIONAL ACCREDITATION BOARD (INAB) is the Irish body for the accreditation of organisations including laboratories.

Laboratory accreditation is available to testing and calibration facilities operated by manufacturing organisations, government departments, educational institutions and commercial testing/calibration services. Indeed, any organisation involved in testing, measurement or calibration in any area of technology can seek accreditation for the work it is undertaking.

Each accredited laboratory has been assessed by skilled specialist assessors and found to meet criteria which are in compliance with ISO/IEC 17025 or ISO/IEC 15189 (medical laboratories). Frequent audits, together with periodic inter-laboratory test programmes, ensure that these standards of operation are maintained.

## Testing and Calibration Categories:

- Category A:** Permanent laboratory calibration and testing where the laboratory is erected on a fixed location for a period expected to be greater than three years.
- Category B:** Site calibration and testing that is performed by staff sent out on site by a permanent laboratory that is accredited by the Irish National Accreditation Board.
- Category C:** Site calibration and testing that is performed in a site/mobile laboratory or by staff sent out by such a laboratory, the operation of which is the responsibility of a permanent laboratory accredited by the Irish National Accreditation Board.
- Category D:** Site calibration and testing that is performed on site by individuals and organisations that do not have a permanent calibration/testing laboratory. Testing may be performed using
- (a) portable test equipment
  - (b) a site laboratory
  - (c) a mobile laboratory or
  - (d) equipment from a mobile or site laboratory

## Standard Specification or Test Procedure Used:

The standard specification or test procedure that is accredited is the issue that is current on the date of the most recent visit, unless otherwise stated.

## Glossary of Terms

### Facilities:

- Public calibration/testing service:** Commercial operations which actively seek work from others.
- Conditionally available for public calibration/testing:** Established for another primary purpose but, more commonly than not, is available for outside work.
- Normally not available for public calibration/testing:** Unavailable for public calibration/testing more often than not.

Laboratory users wishing to obtain assurance that calibration or test results are reliable and carried out to the Irish National Accreditation Board criteria should insist on receiving an accredited calibration certificate or test report. Users should contact the laboratory directly to ensure that this scope of accreditation is current. INAB will on request verify the status and scope.

# Scope of Accreditation



## Adelaide & Meath Hospital (Inc NCH) Blood Transfusion Testing Laboratory

Permanent Laboratory:  
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
1020 Immunohaematology .01 Red Blood Cells	Blood grouping (ABO & RhD Typing).	Manual method Biovue BT-T-EP-001 Tube BT-T-EP-002
.21 Red Blood Cells	Blood grouping (ABO & Rh(D) Typing).	Automated method using the Biovue/Autovue BT-T-EP-004
1020 Immunohaematology .03 Plasma	Antibody Screening	Manual / Biovue BT-T-EP-010 BT-T-EP-004 Automated method using Autovue tubes BT-T-EP-012
.04 Plasma	Antibody Identification	Manual method using Biovue BT-T-EP-014
Red Blood Cells	Confirming ABO and RH(D) group and donor units	Manual method using Autovue Biovue (manual) tubes BT-T-EP-025

# Scope of Accreditation



## Adelaide & Meath Hospital Inc (NCH) Blood Transfusion Testing Laboratory

Permanent Laboratory:  
Category A

INAB Classification number (P9) Materials/products tested	Type of test/properties measured Range of measurement	Standard specifications Equipment/techniques used
1020 Immunohaematology  .05 Plasma	Crossmatch Testing	Manual Method of Crossmatching in Tubes BT-T-EP-006 Biovue BT-T-EP-005 Automated Method BT-T-EP-045
.05 Red Blood Cells	Antigen Typing	Manual Method of Antigen Typing in tubes BT-T-EP-042 Rh + K Phenotyping: DiaMed BT-T-EP-040 Biovue BT-T-EP-041
.05 Red Blood Cells	Direct Coombs Testing	Manual Method of Direct Antiglobulin Testing by DiaMed Cassettes BT-T-EP-030

Note: Adelaide & Meath Hospital Blood Bank has been assessed and is competent to comply with articles 14 and 15 of the EU Directive 2002/98/EC (S.I.No. 360 of 2005 and S.I. No. 547 of 2006)