

Schedule of Accreditation



Organisation Name	Cruinn Diagnostics Limited T/A CruinnCal
Trading As	
INAB Reg No	279C
Contact Name	Brian Ronan
Address	Unit 5B & 6B Hume Center, Hume Center, Park West Industrial Estate, Nangor Road, Dublin, D12
Contact Phone No	01 6297400
Email	info@cruinn.ie
Website	http://www.cruinn.ie
Accreditation Standard	ISO 17025 C
Date Initially Awarded	15/02/2011
Scope Classification	Metrology
Services available to the public ¹	Yes

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered	
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)	
Name	Address
1 Head Office	Unit 5A & 6B,, Hume Centre, Park West Industrial Estate, Dublin, D12

Scope of Accreditation

Head Office

Metrology

Category: A

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
104 Volume - .02 Special laboratory volumetric apparatus	Multi Channel pipettes	1 µl to 5 µl	±0.09 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-020	Multi channel fixed and variable volume pipettes	
		1,000 µl to 2,500 µl	±8.7 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-028	Multi channel fixed and variable volume pipettes	
		10 µl to 20 µl	±0.16 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-022	Multi channel fixed and variable volume pipettes	
		100 µl to 200 µl	±0.8 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-025	Multi channel fixed and variable volume pipettes	
		2,500 µl to 5,000 µl	±22 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-029	Multi channel fixed and variable volume pipettes	
		20 µl to 50 µl	±0.27 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-023	Multi channel fixed and variable volume pipettes	
		200 µl to 500 µl	±2.5 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-026	Multi channel fixed and variable volume pipettes	
		5 µl to 10 µl	±0.11 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-021	Multi channel fixed and variable volume pipettes	
		5,000 µl to 10,000 µl	±75 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-030	Multi channel fixed and variable volume pipettes	

		50 µl to 100 µl	±0.61 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 024	Multi channel fixed and variable volume pipettes	
		500 µl to 1,000 µl	±4.3 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 027	Multi channel fixed and variable volume pipettes	
	Single channel pipettes	.02 µl to 1 µl	±0.09 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 020	Single channel fixed and variable volume pipettes	
		1 µl to 5 µl	± 0.09 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 021	Single channel fixed and variable volume pipettes	
		1,000 µl to 2,500 µl	±8.7 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 029	Single channel fixed and variable volume pipettes	
		10 µl < to 20 µl	±0.16 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 023	Single channel fixed and variable volume pipettes	
		100 µl to 200 µl	±0.8 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 026	Single channel fixed and variable volume pipettes	
		2,500 µl to 5,000 µl	±22 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 030	Single channel fixed and variable volume pipettes	
		20 µl to 50 µl	±0.27 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 024	Single channel fixed and variable volume pipettes	
		200 µl to 500 µl	±2.5 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 027	Single channel fixed and variable volume pipettes	
		5 µl to 10 µl	±0.11 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 022	Single channel fixed and variable volume pipettes	
		5,000 µl v 10,000 µl	±75 µl	ISO 8655 + inhouse procedures CALB- 015, and CALB- 031	Single channel fixed and variable volume pipettes	
		50 µl to 100 µl	±0.61 µl	ISO 8655 + inhouse procedures CALB-	Single channel fixed and variable volume	

				015, and CALB-025	pipettes	
		500 µl to 1,000 µl	±4.3 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-028	Single channel fixed and variable volume pipettes	
104 Volume - .99 Other equipment	Dispensers	1 µl to 5 µl	±0.09 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		1,000 µl to 2,500 µl	±8.7 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		10 µl to 20 µl	±0.16 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		100 µl to 200 µl	±0.80 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		2,500 µl to 5,000 µl	±22 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		20 µl to 50 µl	±0.27 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		200 µl to 500 µl	±2.5 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		5 µl to 10 µl	±0.11 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		5,000 µl to 10,000	±75 µl	Gravimetric	dispensor	

		μl		method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	piston and/or plunger operated volumetric apparatus	
		50 μl to 100 μl	$\pm 0.61 \mu\text{l}$	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		500 μl to 1,000 μl	$\pm 4.3 \mu\text{l}$	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	

Metrology

Category: B

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
101 Mass - .01 Precision laboratory balances	Precision Laboratory Balances	1 mg to 5 g	±0.024 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		100 g to 200 g	±0.27 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		20 g to 50 g	±0.074 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		200 g to 500 g	±0.73 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		50 g to 100 g	±0.14 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		500 g to 1200 g	±2.8 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
		5g to 20g	±0.045mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Precision Laboratory Balances	
101 Mass - .02 Industrial balances	Industrial Balances	1 g to 100 g	±0.21 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
		100 g to 200 g	±0.46 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
		1000 g to 2000 g	±24 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
		200 g to 600 g	±2.8 mg	Documented in-house procedure CALB 029, based	Industrial Balances	

				on EURAMET cg-18		
		2000 g to 5000 g	±25 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
		5000 g to 11000 g	±277 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
		600 g to 1000 g	±3.1 mg	Documented in-house procedure CALB 029, based on EURAMET cg-18	Industrial Balances	
104 Volume - .02 Special laboratory volumetric apparatus	Multi Channel pipettes	1 µl to 5 µl	±0.09 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-020	Multi channel fixed and variable volume pipettes	
		1,000 µl to 2,500 µl	±8.7 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-027	Multi channel fixed and variable volume pipettes	
		10 µl to 20 µl	±0.16 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-021	Multi channel fixed and variable volume pipettes	
		100 µl to 200 µl	±0.8 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-024	Multi channel fixed and variable volume pipettes	
		2,500 µl to 5,000 µl	±22 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-028	Multi channel fixed and variable volume pipettes	
		20 µl to 50 µl	±0.27 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-022	Multi channel fixed and variable volume pipettes	
		200 µl to 500 µl	±2.5 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-025	Multi channel fixed and variable volume pipettes	
		5 µl to 10 µl	±0.11 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-020	Multi channel fixed and variable volume pipettes	
		5,000 µl to 10,000 µl	±75 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-029	Multi channel fixed and variable volume pipettes	
		50 µl to 100 µl	±0.61 µl	ISO 8655 +	Multi channel	

				inhouse procedures CALB-015, and CALB-023	fixed and variable volume pipettes	
		500 µl to 1,000 µl	±4.3 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-026	Multi channel fixed and variable volume pipettes	
Single channel pipettes		.02 µl to 1 µl	±0.09 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-020	Single channel fixed and variable volume pipettes	
		1 µl to 5 µl	±0.09 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-020	Single channel fixed and variable volume pipettes	
		1,000 µl to 2,500 µl	±8.7 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-028	Single channel fixed and variable volume pipettes	
		10 µl to 20 µl	±0.16 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-022	Single channel fixed and variable volume pipettes	
		100 µl to 200 µl	±0.8 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-025	Single channel fixed and variable volume pipettes	
		2,500 µl to 5,000 µl	±22 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-029	Single channel fixed and variable volume pipettes	
		20 µl to 50 µl	±0.27 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-023	Single channel fixed and variable volume pipettes	
		200 µl to 500 µl	±2.5 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-026	Single channel fixed and variable volume pipettes	
		5 µl to 10 µl	±0.11 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-021	Single channel fixed and variable volume pipettes	
		5,000 µl to 10,000 µl	±75 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-030	Single channel fixed and variable volume pipettes	
		50 µl to 100 µl	±0.61 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-	Single channel fixed and variable volume pipettes	

				024		
		500 µl to 1,000 µl	±4.3 µl	ISO 8655 + inhouse procedures CALB-015, and CALB-027	Single channel fixed and variable volume pipettes	
104 Volume - .99 Other equipment	Dispensers	1 µl to 5 µl	±0.09 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		1,000 µl to 2,500 µl	±8.7 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		10 µl to 20 µl	±0.16 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		100 µl to 200 µl	±0.80 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		2,500 µl to 5,000 µl	±22 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		20 µl to 50 µl	±0.27 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		200 µl to 500 µl	±2.5 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		5 µl to 10 µl	±0.11 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		5,000 µl to 10,000 µl	±75 µl	Gravimetric method based on	dispensor piston and/or	

				ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	plunger operated volumetric apparatus	
		50 µl to 100 µl	±0.61 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
		500 µl to 1,000 µl	±4.3 µl	Gravimetric method based on ISO 8655-5:2002 and documented in-house methods CALB-019 and CALB-020.	dispensor piston and/or plunger operated volumetric apparatus	
108 Temperature controlled enclosures - .01 Ovens, furnaces, baths	Ovens, refrigerators, freezers and baths	-20 °C to +130 °C (Using wireless thermistor based data loggers)	±0.12 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Ovens, refrigerators, freezers and baths	
		-90 °C to +130 °C Using PT100 probes	±0.09 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Ovens, refrigerators, freezers and baths	
108 Temperature controlled enclosures - .02 Incubators	Incubators	-20 °C to +130 °C (Using wireless thermistor based data loggers)	±0.12 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Incubators	
		-90 °C to +130 °C Using PT100 probes	±0.09 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Incubators	
108 Temperature controlled enclosures - .03 Autoclaves and sterilising ovens	Autoclaves & Steam Sterilisers	+80 °C to +140 °C	±0.12 °C	In-house SOP CALB 017, CALB 032, EN 60068-3-5:2002, EN60068-3-11:2007	Autoclaves & Steam Sterilisers	
		-20 °C to +130 °C (Using wireless thermistor based data loggers)	±0.12 °C	In-house SOP CALB 017, CALB 032, EN 60068-3-5:2002, EN60068-3-11:2007	Autoclaves & Steam Sterilisers	
		-90 °C to +130 °C Using PT100 probes	±0.09 °C	In-house SOP CALB 017, CALB 032, EN 60068-3-5:2002, EN60068-3-11:2007	Autoclaves & Steam Sterilisers	
108 Temperature controlled enclosures - .04 Industrial freezers	Industrial freezers	-20 °C to +130 °C (Using wireless thermistor based data loggers)	±0.12 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Industrial freezers	
		-80 °C to -20 °C (Using wireless thermistor based data loggers)	±0.34 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3-11:2007	Industrial freezers	

		-90 °C to +130 °C Using PT100 probes	±0.09 °C	In-house SOP CALB 017, EN 60068-3-5:2002, EN60068-3- 11:2007	Industrial freezers	
113 Time - .99 Other	Autoclaves & Steam Sterilisers	0 to 120 minutes	± 0.1 second	In-house SOP CALB 017, CALB 032	Autoclaves & Steam Sterilisers	

Head Office

Metrology

Category: D

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
112 Speed - .01 Centrifuges	Centrifuges	Rotational Speed 500 to 30,000 RPM	±3.2 RPM	Documented inhouse method CALB 030, Calibration of Centrifuges.	Centrifuges	
113 Time - .99 Other	Time	30 to 3600 second	±1 second	Documented inhouse method CALB 030, Calibration of Centrifuges.	Centrifuges	