

Schedule of Accreditation



Organisation Name Laboratory Supplies Ltd
Trading As
INAB Reg No 378C
Contact Name Ken Brereton
Address John F. Kennedy Drive, Naas Road, Dublin 12, Dublin
Contact Phone No 01 4607652
Email ken.brereton@lennox.ie
Website <https://www.lennox.ie>
Accreditation Standard ISO 17025 C
Date Initially Awarded 30/01/2019
Scope Classification Metrology
Services available to the public¹ Yes

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered	
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)	
Name	Address
1 Head Office	John F. Kennedy Drive, Naas Road, Dublin 12, Dublin

Scope of Accreditation

Head Office

Metrology

Category: A

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
107 Temperature measuring equipment - .09 Digital temperature indicator systems	Digital temperature indicator systems	+180 °C to +200 °C -40 °C to +180 °C	0.120 °C 0.029 °C	Documented in house procedure TS.SOP.011.CPM003 Thermometer Calibration		

The Calibration and Measurement Capability expressed in terms of :

*Calibration method
Measurement range
Measurement uncertainty*

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
101 Mass - .01 Precision laboratory balances	Precision laboratory balances	0.001 g to 0.50 g 0.50 g to 5.1 g 5.1 g to 45 g 45 g to 120 g 120 g to 220 g 220 g to 320 g	0.0000070 g 0.000021 g 0.000088 g 0.00017 g 0.00029 g 0.00046 g	Documented in house TS.SOP.009.CPM001 Balance Calibration. Class E2.		
101 Mass - .02 Industrial balances	Industrial balances	320 g to 520 g 520 g to 1500 g 1500 g to 6100 g	0.00071 g 0.0027 g 0.021 g	Documented in house TS.SOP.009.CPM001 Balance Calibration. Class E2		
		6100 g to 14200 g 14200 g to 32100 g 32100 g to 50000 g 50000 g to 64100 g 64100 g to 70000 g	0.038 g 0.21 g 0.47 g 0.61 g 1.4 g	Documented in house TS.SOP.009.CPM001 Balance Calibration. Combined Class E2 / F1 / M1		
108 Temperature controlled enclosures - .01 Ovens, furnaces, baths	Ovens, Furnaces and baths	-40 °C to +200 °C -40 °C to +200 °C	0.16 °C (Pt-100) 0.27 °C (Thermocouples)	Single and Multi-point calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration. Various temperature enclosures	Temperature mapping	
108 Temperature controlled enclosures - .02 Incubators	Incubators	-40 °C to +200 °C -40 °C to +200 °C	0.16 °C (Pt-100) 0.27 °C (Thermocouples)	Single and Multi-point calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration. Temperature mapping.		
108 Temperature controlled enclosures - .03 Autoclaves and sterilising ovens	Autoclaves and sterilising ovens	-40 °C to +200 °C -40 °C to +200 °C	0.16 °C (Pt-100) 0.27 °C (Thermocouples)	Single and Multi-point calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration Various temperature enclosures	Temperature mapping	
108 Temperature controlled enclosures - .04 Industrial freezers	Industrial freezers	-40 °C to +200 °C -40 °C to +200 °C	0.16 °C (Pt-100) 0.27 °C (Thermocouples)	Single and Multi-point calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration. Temperature mapping.		
108 Temperature controlled enclosures - .99 Other	testing of temperature controlled enclosures	-40 °C to +200 °C -40 °C to +200 °C	0.16 °C (Pt-100) 0.27 °C (Thermocouples)	Single and Multi-point calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration Various temperature enclosures.	Temperature mapping	
108 Temperature	Fridges	-40 °C to +200 °C	0.16 °C (Pt-100)	Single and Multi-point		

controlled enclosures - 0.05 Fridges		-40 °C to +200 °C	0.27 °C (Thermocouples)	calibration using in-house procedure TS.SOP.010.CPM002 Temperature Enclosures Calibration. Temperature mapping.		
---	--	-------------------	----------------------------	--	--	--

The Calibration and Measurement Capability expressed in terms of :

Calibration method
Measurement range
Measurement uncertainty