

Schedule of Accreditation



Organisation Name	Mason Technology Ltd
Trading As	
INAB Reg No	043C
Contact Name	Mandy Nolan
Address	228 South Circular Road, Dublin, D08 DX8P
Contact Phone No	01 4534422
Email	mnolan@masontec.ie
Website	http://www.masontechnology.ie
Accreditation Standard	ISO 17025 C
Date Initially Awarded	26/07/1994
Scope Classification	Metrology
Services available to the public ¹	Yes

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered	
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)	
Name	Address
1 Head Office	228 South Circular Road, Dublin, D08 DX8P

Scope of Accreditation

Head Office

Metrology

Category: A

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
101 Mass - .01 Precision laboratory balances	Precision laboratory balances	1 mg - 5 g	0.020 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights
		5 g - 22 g	0.026 mg			OIML Class F1 Weights
		22 g - 100 g	0.051 mg			
		100 g - 600 g	0.14 mg			
		600 g - 1000 g	1.0 mg			
		1000 g - 8100 g	1.4 mg			
		20 g - 600 g	0.60 mg			
		600 g - 1000 g	2.5 mg			
		1 kg - 5 kg	6.0 mg			
		5 kg - 20 kg	10 mg			
		20 kg - 60 kg	65 mg			
101 Mass - .02 Industrial balances	Industrial balances	1 mg - 5 g	0.020 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights
		5 g - 22 g	0.026 mg			OIML Class F1 Weights
		22 g - 100 g	0.051 mg			
		100 g - 600 g	0.14 mg			
		600 g - 1000 g	1.0 mg			
		1000 g - 8100 g	1.4 mg			
		20 g - 600 g	0.60 mg			
		600 g - 1000 g	2.5 mg			
		1 kg - 5 kg	6.0 mg			

		5 kg - 20 kg 20 kg - 60 kg	10 mg 65 mg			
101 Mass - .03 Industrial weighing appliances	Weighing appliances	1 mg - 5 g 5 g - 22 g 22 g - 100 g 100 g - 600 g 600 g - 1000 g 1000 g - 8100 g 20 g – 600 g 600 g - 1000 g 1 kg - 5 kg 5 kg - 20 kg 20 kg - 60 kg	0.020 mg 0.026 mg 0.051 mg 0.14 mg 1.0 mg 1.4 mg 0.60 mg 2.5 mg 6.0 mg 10 mg 65 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights OIML Class F1 Weights
101 Mass - .05 Laboratory mass standards	Laboratory mass standards	1000 g 500 g 200 g 100 g 50 g 20 g 10 g 5 g 2 g 1 g 0.5 g 0.2 g 0.1 g 0.05 g 0.02 g 0.01 g 0.005 g 0.002 g 0.001 g	0.40 mg 0.20 mg 0.075 mg 0.040 mg 0.025 mg 0.020 mg 0.015 mg 0.013 mg 0.010 mg 0.0075 mg 0.0063 mg 0.0050 mg 0.0040 mg 0.0040 mg 0.0033 mg 0.0027 mg 0.0020 mg 0.0020 mg 0.0020 mg	SOP MS1 SOP MS4	E2 class Weights	
		20000 g 10000 g 5000 g 2000 g 1000 g 500 g 200 g 100 g 50 g 20 g	20 mg 10 mg 8.3 mg 2.0 mg 1.0 mg 0.50 mg 0.20 mg 0.12 mg 0.060 mg 0.050 mg	SOP MS1 SOP MS4	F and M Class Weights	

		10 g 5 g 2 g 1 g 0.5 g 0.2 g 0.1 g 0.05 g 0.02 g 0.01 g 0.005 g 0.002 g 0.001 g	0.040 mg 0.030 mg 0.024 mg 0.020 mg 0.016 mg 0.012 mg 0.010 mg 0.0080 mg 0.0060 mg 0.0050 mg 0.0040 mg 0.0040 mg 0.0048 mg			
104 Volume - .02 Special laboratory volumetric apparatus	Multi Channel Pipettes	0.1 to 5µl 5 to 20µl 20 to 100µl 100 to 200µl 200 to 500µl 500 to 1,000µl 1,000 to 2,000µl 2,000 to 5,000µl 5,000 to 10,000µl	0.11 µl 0.20 µl 0.25 µl 0.30 µl 0.50 µl 1.0 µl 4.0 µl 5.0 µl 10 µl	SOP PIP1 SOP PIP2	Multi Channel Pipettes	Multi channel fixed and variable volume pipettes
	Single Channel Pipettes	0.1 to 5µl 5 to 20µl 20 to 100µl 100 to 200µl 200 to 500µl 500 to 1,000µl 1,000 to 2,000µl 2,000 to 5,000µl 5,000 to 10,000µl 10ml to 50ml burettes, dispensers 50ml to 100ml burettes, dispensers	0.11 µl 0.20 µl 0.25 µl 0.30 µl 0.50 µl 1.0 µl 4.0 µl 5.0 µl 10 µl 40 µl 60 µl	SOP PIP1 SOP PIP2	Single Channel Pipettes	Single channel fixed and variable volume pipettes
107 Temperature measuring equipment - .09 Digital temperature indicator systems	Temperature - All Other Probes	Permanent -86°C to 0°C 0°C to +125°C +125°C to +300°C	Permanent 0.084°C + Unit Resolution 0.048°C + Unit Resolution 0.115°C + Unit Resolution	SOP TS1	Thermometers/ Digital Indicators	

		Permanent (Temperature Sensors Incorporated into Humidity Instruments) 0°C to +15°C +15°C to + 30°C +30°C to +60°C	0.14°C+ Unit Resolution 0.11°C + Unit Resolution 0.14°C + Unit Resolution	SOP RH1	Temperature/ Humidity Devices	
	Temperature - Thermocouples	Permanent -86°C to 0 °C 0°C to +125°C +125°C to +300°C	0.22°C + Unit Resolution 0.21°C + Unit Resolution 0.24°C + Unit Resolution	SOP TS1	Thermometers/ Digital Indicators	
108 Temperature controlled enclosures - .01 Ovens, furnaces, baths	Temperature	-80°C to 0°C (WS) 0C° to +140°C (WS) +140°C to +300°C (WS) -80°C to +75°C (Veriteqs)	0.48°C 0.18°C 0.48°C 0.32°C	SOP T1 SOP T2	Ovens, Baths, Blocks, Temperature Controlled Rooms	
108 Temperature controlled enclosures - .02 Incubators	Temperature	-10°C to +100°C (WS) -10°C to +75°C (Veriteqs)	0.18°C 0.32°C	SOP T3	Incubators	
108 Temperature controlled enclosures - .03 Autoclaves and sterilising ovens	Temperature/ Time	Autoclaves: +80°C to +140°C 5 mins. to 24 hrs. time interval Sterilising Ovens: +80°C to +140°C 5 mins. to 24 hrs. time interval	0.18 °C 1.6 secs per 1 hour 0.18 °C 1.6 secs per 1 hour	SOP T6 SOP T1	Autoclaves Sterilising Ovens	

108 Temperature controlled enclosures - .04 Industrial freezers	Temperature	-80°C to 0°C (WS) -80°C to 0°C (Veriteqs)	0.48°C 0.32°C	SOP T1	Freezers, Temperature Controlled Rooms	
108 Temperature controlled enclosures - 0.05 Fridges	Temperature	0°C to +10°C (WS) 0°C to +10°C (Veriteqs)	0.48°C 0.32°C	SOP T1	Fridges, Temperature Controlled Rooms	
112 Speed - .01 Centrifuges	Speed	0 to 599.9 rpm 599.9 to 9,999.9 rpm 10,000 to 50,000 rpm Elapsed time 1 second to 1 hour	2.2 rpm 2.9 rpm 13 rpm 0.18 seconds	SOP SMD1	Centrifuges	
116 Hygrometry - .01 Humidity testing device	Humidity/ Temperature Devices	5% to 95%rH in the range of 0°C to 60°C 0°C to +15°C 5%rH to 90%rH: +15°C to +45°C 5%rH to 95%rH: +45°C to +60°C 5%rH to 90%rH Temperature sensors incorporated in humidity instruments 0°C to +15°C +15°C to +30°C +30°C to +60°C	0.22%rH to 1.9%rH 0.17%rH to 1.2%rH 0.15%rH to 1.0%rH 0.14°C + Unit Resolution 0.11°C + Unit Resolution 0.14°C + Unit Resolution	SOP RH1	Temperature/Humidity Devices	
116 Hygrometry - .02 Environmental Chambers	Temperature/ Humidity	10°C to 60°C 20%rH to 85%rH	0.32°C 3.3%rH	SOP T7	Stability, Climatic, Environmental Chambers	

Mass calibration laboratory: nominal temperature for calibration work: 20±1°C for E2 Class weights, 20±2°C for F Class weights, 22.5±3.5°C for M Class weights
Balance calibration laboratory: nominal temperature for calibration work: 20±5°C
Speed device calibration laboratory: nominal temperature for calibration work:

$20\pm 15^{\circ}\text{C}$

Temperature calibration laboratory: nominal temperature for calibration work:

$20\pm 15^{\circ}\text{C}$ (probes and thermometers)

Temperature calibration laboratory: nominal temperature for calibration work:

$20\pm 15^{\circ}\text{C}$ (enclosures)

Humidity calibration laboratory: nominal temperature for calibration work:

$20\pm 10^{\circ}\text{C}$ (hygrometry measuring devices)

Humidity calibration laboratory: nominal temperature for calibration work:

$20\pm 15^{\circ}\text{C}$ (stability, climatic, environmental chambers)

Pipette Calibration laboratory: nominal temperature for calibration work : $20\pm 1^{\circ}\text{C}$

Calibration Measurement Capability (CMC) is expressed in terms of the following parameters:

- Measurand or reference material*
- Calibration or measurement method or procedure and type of instrument or material calibrated/measured*
- Measurement range and additional parameters where applicable*
- Measurement uncertainty.*

Measurement uncertainty shall be reported in compliance with EA 4/02

“Expression of the Uncertainty of Measurement in Calibration”.

In accordance with INAB policy, uncertainties are calculated for an estimated confidence level of not less than 95%.

Head Office

Metrology

Category: B

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Calibration and measurement capability (CMC)	Std. ref/SOP	Products	Remarks
101 Mass - .01 Precision laboratory balances	Precision laboratory balances	1 mg - 5 g 5 g - 22 g 22 g - 100 g 100 g - 600 g 600 g - 1000 g 1000 g - 8100 g 20 g – 600 g 600 g - 1000 g 1 kg - 5 kg 5 kg - 20 kg 20 kg - 60 kg	0.020 mg 0.026 mg 0.051 mg 0.14 mg 1.0 mg 1.4 mg 0.60 mg 2.5 mg 6.0 mg 10 mg 65 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights OIML Class F1 Weights
101 Mass - .02 Industrial balances	Industrial balances	1 mg - 5 g 5 g - 22 g 22 g - 100 g 100 g - 600 g 600 g - 1000 g 1000 g - 8100 g 20 g – 600 g 600 g - 1000 g 1 kg - 5 kg 5 kg - 20 kg 20 kg - 60 kg	0.020 mg 0.026 mg 0.051 mg 0.14 mg 1.0 mg 1.4 mg 0.60 mg 2.5 mg 6.0 mg 10 mg 65 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights OIML Class F1 Weights
101 Mass - .03 Industrial weighing appliances	Weighing appliances	1 mg - 5 g 5 g - 22 g 22 g - 100 g 100 g - 600 g 600 g - 1000 g 1000 g - 8100 g	0.020 mg 0.026 mg 0.051 mg 0.14 mg 1.0 mg 1.4 mg	SOP W1 SOP W2	Balances and Industrial Scales	OIML Class E2 Weights

		20 g – 600 g 600 g - 1000 g 1 kg - 5 kg 5 kg - 20 kg 20 kg - 60 kg	0.60 mg 2.5 mg 6.0 mg 10 mg 65 mg			OIML Class F1 Weights
104 Volume - .02 Special laboratory volumetric apparatus	Multi Channel Pipettes	0.2 to 5µl 5 to 20µl 20 to 100µl 100 to 200µl 200 to 500µl 500 to 1,000µl 1,000 to 2,000µl 2,000 to 5,000µl 5,000 to 10,000µl	0.11 µl 0.20 µl 0.25 µl 0.30 µl 0.50 µl 1.0 µl 4.0 µl 5.0 µl 10 µl	SOP PIP1 SOP PIP2	Multi Channel Pipettes	Multi channel fixed and variable volume pipettes
	Single Channel Pipettes	0.2 to 5µl 5 to 20µl 20 to 100µl 100 to 200µl 200 to 500µl 500 to 1,000µl 1,000 to 2,000µl 2,000 to 5,000µl 5,000 to 10,000µl 10ml to 50ml burettes, dispensers 50ml to 100ml burettes, dispensers	0.11 µl 0.20 µl 0.25 µl 0.30 µl 0.50 µl 1.0 µl 4.0 µl 5.0 µl 10 µl 40 µl 60 µl	SOP PIP1 SOP PIP2	Single Channel Pipettes	Single channel fixed and variable volume pipettes
107 Temperature measuring equipment - .09 Digital temperature indicator systems	Temperature - All Other Probes	Site -80°C to -30°C -30 °C +140°C +140°C to +250°C	Site 0.55°C + Unit Resolution 0.22°C + Unit Resolution 0.33°C + Unit Resolution	SOP TS1	Thermometers/ Digital Indicators	
	Temperature - Thermocouples	Site -80°C to -30°C -30C° to +140°C +140°C to +250°C	Site 0.59°C + Unit Resolution 0.30°C + Unit Resolution 0.39°C + Unit Resolution	SOP TS1	Thermometers/ Digital Indicators	

108 Temperature controlled enclosures - .01 Ovens, furnaces, baths	Temperature	Site -80°C to 0°C (WS) 0°C to +140°C (WS) +140°C to +300°C (WS) -80°C to +75°C (Veriteqs)	0.48°C 0.18°C 0.48°C 0.32°C	SOP T1 SOP T2	Ovens, Baths, Blocks, Temperature Controlled Rooms
108 Temperature controlled enclosures - .02 Incubators		-10°C to +100°C (WS) -10°C to +75°C (Veriteqs)	0.18°C 0.32°C	SOP T3	Incubators
108 Temperature controlled enclosures - .03 Autoclaves and sterilising ovens	Temperature/ Time	Autoclaves: +80°C to +140°C 5 mins. to 24 hrs. time interval	0.18°C 1.6 secs per 1 hour	SOP T6	Autoclaves
		Sterilising Ovens: +80°C to +140°C 5 mins. to 24 hrs. time interval	0.18 °C 1.6 secs per 1 hour	SOP T1	Sterilising Ovens
108 Temperature controlled enclosures - .04 Industrial freezers	Temperature	-80°C to 0°C (WS) -80°C to 0°C (Veriteqs)	0.48°C 0.32°C	SOP T1	Freezers, Temperature Controlled Rooms
108 Temperature controlled enclosures - 0.05 Fridges	Temperature	0°C to +10°C (WS) 0°C to +10°C (Veriteqs)	0.48°C 0.32°C	SOP T1	Fridges, Temperature Controlled Rooms
112 Speed - .01 Centrifuges	Speed	0 to 599.9 rpm 599.9 to 9,999.9 rpm 10,000 to 50,000 rpm	2.2 rpm 2.9 rpm 13 rpm	SOP SMD1	Centrifuges
		Elapsed time 1 second to 1 hour	0.18 seconds		
116 Hygrometry - .02 Environmental Chambers	Temperature/ Humidity	10°C to 60°C 20%rH to 85%rH	0.32°C 3.3%rH	SOP T7	Stability, Climatic, Environmental Chambers

Mass calibration laboratory: nominal temperature for calibration work: 20±1°C for E2 Class weights, 20±2°C for F Class weights, 22.5±3.5°C for M Class weights
Balance calibration laboratory: nominal temperature for calibration work: 20±5°C
Speed device calibration laboratory: nominal temperature for calibration work: 20±15°C

*Temperature calibration laboratory: nominal temperature for calibration work:
20±15°C (probes and thermometers)*

*Temperature calibration laboratory: nominal temperature for calibration work:
20±15°C (enclosures)*

*Humidity calibration laboratory: nominal temperature for calibration work:
20±10°C (hygrometry measuring devices)*

*Humidity calibration laboratory: nominal temperature for calibration work:
20±15°C (stability, climatic, environmental chambers)*

Pipette Calibration laboratory: nominal temperature for calibration work : 20±1°C

*Calibration Measurement Capability (CMC) is expressed in terms of the
following parameters:*

- Measurand or reference material*
- Calibration or measurement method or procedure and type of instrument or material calibrated/measured*
- Measurement range and additional parameters where applicable*
- Measurement uncertainty.*

*Measurement uncertainty shall be reported in compliance with EA 4/02
“Expression of the Uncertainty of Measurement in Calibration”.*

*In accordance with INAB policy, uncertainties are calculated for an estimated
confidence level of not less than 95%.*