

Schedule of Accreditation



Organisation Name Bantry General Hospital
Trading As
INAB Reg No 243MT
Contact Name Hugh Brennan
Address Blood Transfusion Section, Laboratory Department,
Bantry, Cork
Contact Phone No 00353 27 52910
Email hugh.brennan@hse.ie
Website
Accreditation Standard ISO 15189
Date Initially Awarded 15/09/2009
Scope Classification Blood Transfusion Science

Services available to the public¹

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered		
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)		
	Name	Address
1	Head Office	Blood Transfusion Section, Laboratory Department, Bantry, Cork, P75 DX93

Scope of Accreditation

Head Office

Blood Transfusion Science

Category: A

Medical pathology field - Test	Test/assay	Specimen Type	Equipment/Technique	Method (CE/Non-CE/In house developed/based on standard method)	Range of measurement	Std. ref & SOP
1020 Transfusion science - .01 Blood grouping including ABO, Rh(D) and other antigens by manual methods	ABO & Rh(D) typing	Blood (EDTA / Serum)	Manual / Tube NISS & Monoclonal Antisera	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/005
1020 Transfusion science - .02 Blood grouping including ABO, Rh(D) and other antigens by automated methods			DiaMed Gelstation/ Tube NISS & Monoclonal Antisera	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/005 BGH-QA-SOP-0150
1020 Transfusion science - .03 Blood group antibody screen	ABO & Rh(D) typing / Antibody Screen		Manual / DiaMed Gel Card Method by direct and indirect agglutination	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/005
1020 Transfusion science - .04 Identification of blood group antibodies	Antibody Identification	EDTA Plasma / Serum	DiaMed Gelstation/ Gel Card (IAT and Direct Agglutination)	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/009 BGH-QA-SOP-0150
1020 Transfusion science - .05 Cross match compatible donor units	Crossmatch		Manual / DiaMed Gel Card Method by direct and indirect agglutination	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/007
1020 Transfusion science - .06 Red cell phenotyping	Phenotype	Blood (EDTA / Serum)	Manual / DiaMed Gel Card Method by direct and indirect agglutination	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/012
1020 Transfusion science - .09 Direct antiglobulin test	Direct Antiglobulin Test		Manual / DiaMed Gel Card Method by indirect agglutination	CE / Based on standard method	Positive / Negative	BGH/BTR/SOP/019

The hospital blood bank has been assessed and is competent to comply with Articles 14 and 15 of the EU Directive 2002/98/EC (S.I. 360/2005 and S.I. 547/2006)