

Schedule of Accreditation



Organisation Name	Sligo Regional Hospital (Blood Transfusion)
INAB Reg No	198MT
Contact Name	Sonia Gilmartin
Address	Blood Transfusion Department, The Mall, Sligo, Sligo
Contact Phone No	071-9174562
Email	sonia.gilmartin@hse.ie
Website	
Accreditation Standard	ISO 15189
Date Initially Awarded	31/05/2011
Scope Classification	Blood Transfusion Science
Services available to the public ¹	No

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered	
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)	
Name	Address
1 Head Office	Blood Transfusion Department, Sligo, F91 H684

Scope of Accreditation

Head Office

Blood Transfusion Science

Category: A

Medical pathology field - Test	Test/assay	Specimen Type	Equipment/Technique	Method (CE/Non-CE/In house developed/based on standard method)	Range of measurement	Std. ref & SOP
1020 Transfusion science - .01 Blood grouping including ABO, Rh(D) and other antigens by manual methods	ABO-DD grouping (Forward and Reverse)	Whole blood Plasma	Ortho BioVue CAT cassettes, Ortho Work Station and Tubes	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-54 (Manual, Ortho Cassettes) PATH-BT-LP-55 (manual, Tube Technique)
1020 Transfusion science - .02 Blood grouping including ABO, Rh(D) and other antigens by automated methods			Ortho BioVue CAT cassettes and Innova analyser	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-85 (automated, Ortho Innova)
1020 Transfusion science - .03 Blood group antibody screen	Antibody Screening		Ortho BioVue CAT cassettes, Ortho Workstation, Innova analyser	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-54 (Manual, Ortho Cassettes) PATH-BT-LP-85 (automated, Ortho innova)
1020 Transfusion science - .04 Identification of blood group antibodies	Antibody identification using panels		Ortho BioVue CAT cassettes, Ortho Workstation, Innova analyser	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-85 (automated, Ortho innova), PATH-BT-LP-102 (manual Ortho), PATH-BT-LP-131 (Investigation of a positive antibody screen),
1020 Transfusion science - .05 Cross match compatible donor units	Cross match compatibility testing		Ortho BioVue CAT cassettes, Ortho Workstation, Innova analyser	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-85 (automated, Ortho innova), PATH-BT-LP-57 (Transfusion reaction), PATH-BT-LP-35 (manual, Ortho), PATH-BT-LP-14 (Selection of blood for crossmatch)
1020 Transfusion science - .06 Red cell	Red cell Phenotyping		Ortho BioVue CAT cassettes, Ortho	CE standard method as per	1-4 positive, 0 negative	PATH-BT-LP-85 (automated,

phenotyping			Workstation, Innova analyser	manufacturer		Ortho innova), PATH-BT-LP-37 (manual Antigen)
1020 Transfusion science - .09 Direct antiglobulin test	DAT		Ortho BioVue CAT cassettes, Ortho Workstation, Innova analyser	CE standard method as per manufacturer	1-4 positive, 0 negative	PATH-BT-LP-27 (manual Ortho), PATH-BT-LP-85 (automated Ortho Innova), PATH-BT-LP-133 Manual DAT by tube