

# Schedule of Accreditation



Organisation Name VWR International Ltd  
Trading As  
INAB Reg No 126T  
Contact Name Niamh Kinsella  
Address Orion Business Campus, Northwest Business Park,  
Ballycoolin, Dublin, D15  
Contact Phone No 1-8822222  
Email Niamh.Kinsella@vwr.com  
Website <https://ie.vwr.com>  
Accreditation Standard ISO 17025 T  
Date Initially Awarded 03/04/2002  
Scope Classification Mechanical testing

Services available to the public<sup>1</sup>

<sup>1</sup> Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered	
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)	
Name	Address
1 Head Office	Orion Business Campus, Northwest Business Park, Ballycoolin, Dublin, D15

# Scope of Accreditation

## Head Office

### Mechanical Testing

Category: A

Product categories - Tests	Test detail	Product detail	Range of Measurement	Equipment/Technique	Std. Ref & SOP	
1169 Controlled environments - .99 Other tests	Airflow Measurements from Air Samplers	Air Samplers		Testing to manufacturer specification as given	CAL/PRO/22	
	Spectral transmittance, wavelength, resolution and stray light	Spectrophotometers	0A to 3A, 190nm to 700nm	Liquid filters	Cal/Pro/21 based on USP 38 (857) Ultraviolet visible spectroscope pages 664 to 666	

Mechanical Testing

Category: B

Product categories - Tests	Test detail	Product detail	Range of Measurement	Equipment/Technique	Std. Ref & SOP	
1169 Controlled environments - .01 Physical tests	Particle counting, airflow velocity, filter integrity	Laminar Flow workstations	0.25m/s to 2m/s	As specified and in accordance with EN14644	EN14644, PD6609-2000. Cal/Pro/14	
1169 Controlled environments - .02 Chemical tests	Filter and seal integrity testing, Operator protection test ( KI DISCUS), airflow visualisation, airflow velocity	Microbiological safety cabinets Class 1/2/3	0.25m/s to 2m/s	As specified and in accordance with EN12469	EN12469-2000. Cal/Pro/10	
1169 Controlled environments - .99 Other tests	Particle counting, airflow velocity, Filter and seal integrity testing, Operator protection test ( KI DISCUS), airflow visualisation	Air handling systems, Clean Air Devices Clean Zones Local Exhaust Ventilation Devices Chemical Safety Cabinets Re-circulating Filtration Fume Cupboards HEPA Fan Filter Device Isolators Balance enclosures Fan Filter Units	0.25m/s to 2m/s	Testing to manufacturer specification as given	Testing to manufacturer specification as given	
	Spectral transmittance, wavelength, resolution and stray light	Spectrophotometers	0A to 3A, 190nm to 700nm	Liquid filters		