

Schedule of Accreditation



Organisation Name	Endress and Hauser (Ireland) Ltd
INAB Reg No	428C
Contact Name	Joe Bonfield
Address	Block 2, Exchequer House, Embassy Office Park, Kill, Kildare, W91 W866
Contact Phone No	0867877145
Email	joe.bonfield@endress.com
Website	www.ie.endress.com
Accreditation Standard	EN ISO/IEC 17025 C
Standard Version	2017
Date of award of accreditation	11/06/2025
Scope Classification	Metrology
Services available to the public ¹	No

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered		
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)		
	Name	Address
1	Endress + Hauser Cork	Unit4A Fota Business Park, Carrigtwohill, Cork, Ireland, T45 EH50

Scope of Accreditation

Endress + Hauser Cork

Metrology

Category: A

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Expanded Measurement Uncertainty	Std. ref/SOP	Products	Remarks
105 Flow - .11 Liquid meters	Mass Flow Rate	100 kg/h to 60,000 kg/h	0.13% of reading	LAB_SOP_001 - Standard Operating Procedure for Laboratory Flow Calibrations	Coriolis Flow Meters	Digital or analogue output Master meter method Flying stop & start
	Volume Flow Rate	100 l/h to 60,000 l/h	0.16% of reading	LAB_SOP_001 - Standard Operating Procedure for Laboratory Flow Calibrations	Coriolis Flow Meters	Digital or analogue output Master meter method Flying stop & start
			0.18% of reading	LAB_SOP_001 - Standard Operating Procedure for Laboratory Flow Calibrations	Electromagnetic Flow Meters	Digital or analogue output Master meter method Flying stop & start
			1.43% of reading	LAB_SOP_001 - Standard	Vortex Flow Meters	Digital or analogue output

				Operating Procedure for Laboratory Flow Calibrations		Master meter method Flying stop & start
--	--	--	--	---	--	--