# Schedule of Accreditation



Organisation Name	Medical Supply Co Ltd
Trading As	Medical Supply Company
INAB Reg No	262C
Contact Name	Thomas Holmes
Address	Damastown, Mulhuddart, Dublin, D15
Contact Phone No	01 822 4222
Email	tholmes@medical-supply.ie
Website	
Accreditation Standard	EN ISO/IEC 17025 C
Standard Version	2017
Date of award of accreditation	01/06/2010
Scope Classification	Metrology
Services available to the public <sup>1</sup>	Yes

<sup>1</sup> Refer to document on interpreting INAB Scopes of Accreditation

	Sites from which accredited services are delivered								
	(the detail of the accredited services delivered at each site are on the Scope of Accreditation)								
	Name Address								
1	Head Office	Damastown, Mulhuddart, Dublin, D15							

## Scope of Accreditation

#### **Head Office**

#### Metrology

Category: A

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Expanded Measurement Uncertainty	Std. ref/SOP	Products	Remarks
107 Temperature measuring equipment09 Digital temperature indicator systems	0 °C to 37 °C	0 °C to 37 °C	0.06°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
			0.07°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
			0.09 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
	0°C to 85°C	0°C to 85°C	0.11°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Thermobuttons Min / Max Sensors	Single and Multi-point Calibration of temperature sensors
	100°C to 140°C	100°C to 140°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
	140°C	140°C	0.10°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors

		le la construcción de la			
-20 °C to 0 °C	-20 °C to 0 °C	0.37 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
-20 °C to 5 C	-20 °C to 5 C	0.23°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
-20°C to 0°C	-20°C to 0°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
		0.13°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Thermobuttons Min / Max Sensors	Single and Multi-point Calibration of temperature sensors
37 °C to 60 °C	37 °C to 60 °C	0.09°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
37 °C to 70 °C	37 °C to 70 °C	0.10 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
37°C to 60°C	37°C to 60°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
-40 °C to -20 °C	-40 °C to -20 °C	0.31°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
-40°C to -20°C	-40°C to -20°C	0.06°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
5 °C to 37 °C	5 °C to 37 °C	0.16°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
60 ° to 140 °C	60 ° to 140 °C	0.09°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
60°C to 100°C	60°C to 100°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
-90 °C to -20 °C	-90 °C to -20 °C	0.21 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
-90 °C to -40 °C	-90 °C to -40 °C	0.46°C	SOP# 18, SOP# 23, SOP#31,	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature

				SOP# 34		sensors.
	-90°C to -40°C	-90°C to -40°C	0.10°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
108 Temperature controlled enclosures01 Ovens, furnaces, baths	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	100°C to 140°C 100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors	
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors

					1	
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
108 Temperature controlled enclosures02 Incubators	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	100°C to 140°C	100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using

						1
						thermistors
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
108 Temperature controlled enclosures03 Autoclaves and sterilising ovens	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	100°C to 140°C	0°C 100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
140°C	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled

						enclosures using thermistors
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
08 Temperature controlled enclosures04 ndustrial freezers	-20°C to 0°C	-20°C to 0°C	0.12°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezer
	-40°C to -20°C	-40°C to -20°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various freezers from - 40°C to -20°C	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezers
	-90°C to -40°C	-90°C to -40°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various freezers from - 90°C to -40°C	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezer

Measurement uncertainty.
Measurement uncertainty shall be reported in compliance with EA 4/02 "Evaluation of the Uncertainty of Measurement in Calibration".
In accordance with INAB policy, uncertainties are calculated for an estimated confidence level of not less than 95%.

### **Head Office**

### Metrology

#### Category: B

Metrology field - Calibrated Device Type	Measured quantity	Calibration range	Expanded Measurement Uncertainty	Std. ref/SOP	Products	Remarks
107 Temperature measuring equipment09 Digital temperature indicator systems	0 °C to 37 °C	0 °C to 37 °C	0.06°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
			0.07°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
			0.09 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
	0°C to 85°C	0°C to 85°C	0.11°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Thermobuttons Min / Max Sensors	Single and Multi-point Calibration of temperature sensors
	100°C to 140°C	100°C to 140°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
	140°C	140°C	0.10°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
	-20 °C to 0 °C	-20 °C to 0 °C	0.37 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
	-20 °C to 5 C	-20 °C to 5 C	0.23°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
	-20°C to 0°C	-20°C to 0°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors

			0.13°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Thermobuttons Min / Max Sensors	Single and Multi-point Calibration of temperature sensors
	37 °C to 60 °C	37 °C to 60 °C	0.09°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint
	37 °C to 70 °C	37 °C to 70 °C	0.10 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
	37°C to 60°C	37°C to 60°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
	-40 °C to -20 °C	-40 °C to -20 °C	0.31°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
	-40°C to -20°C	-40°C to -20°C	0.06°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
	5 °C to 37 °C	5 °C to 37 °C	0.16°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
	60 ° to 140 °C	60 ° to 140 °C	0.09°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
	60°C to 100°C	60°C to 100°C	0.08°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	Resistance thermistors	Single and Multi-point Calibration of temperature sensors
	-90 °C to -20 °C	-90 °C to -20 °C	0.21 °C	SOP 47, SOP 48	Handheld Thermometer with smart probes	Comparison calibration of temperature sensors.
	-90 °C to -40 °C	-90 °C to -40 °C	0.46°C	SOP# 18, SOP# 23, SOP#31, SOP# 34	PT1000 - Resistance Temperature Detectors	Single and multipoint calibration of temperature sensors.
	-90°C to -40°C	-90°C to -40°C	0.10°C	SOP # 18, SOP # 23, SOP # 31, SOP # 34	T-Type Thermocouples K - Type Thermocouples	Single and Multi-point Calibration of temperature sensors
108 Temperature controlled enclosures01 Ovens, furnaces, baths	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of

						temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controllec enclosures using thermistors
	100°C to 140°C	100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
108 Temperature controlled enclosures02	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP # 22, SOP # 40,	Various	Single and Multi-point temperature mapping

Incubators				SOP # 41		and calibration of temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	100°C to 140°C	100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
108 Temperature	0°C to 25°C	0°C to 25°C	0.10°C	SOP # 19, SOP #	Various	Single and Multi-poir

controlled enclosures03 Autoclaves and sterilising ovens				22, SOP # 40, SOP # 41		temperature mapping and calibration of temperature controlled enclosures using thermocouples.
			0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	100°C to 140°C	100°C to 140°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
			0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	140°C	140°C	0.14°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples.
	25°C to 50°C	25°C to 50°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors
	50°C to 100°C	50°C to 100°C	0.11°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermistors

108 Temperature controlled enclosures04 Industrial freezers	-20°C to 0°C	-20°C to 0°C	0.12°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezers
	-40°C to -20°C	-40°C to -20°C	0.15°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various freezers from - 40°C to -20°C	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezers
	-90°C to -40°C	-90°C to -40°C	0.13°C	SOP # 19, SOP # 22, SOP # 40, SOP # 41	Various freezers from - 90°C to -40°C	Single and Multi-point temperature mapping and calibration of temperature controlled enclosures using thermocouples. Includes refrigerators and freezers
Calibration Measurement C Measurand or reference Calibration or measurem Measurement range and Measurement uncertainty Measurement uncertainty s In accordance with INAB po	material ent method or procec additional parameter /. hall be reported in co	ure and type of instrum s where applicable mpliance with EA 4/02	ent or material calib "Evaluation of the U	rated/measured ncertainty of Measurement	in Calibration".	