

Schedule of Accreditation



Organisation Name	Mater Private Network Cork
Trading As	
INAB Reg No	416MT
Contact Name	Louise O'Callaghan
Address	City Gate, Mahon, Cork
Contact Phone No	0216013411
Email	louise.ocallaghan@materprivate.ie
Website	
Accreditation Standard	EN ISO 15189
Standard Version	2022
Date of award of accreditation	17/04/2024
Scope Classification	Blood Transfusion Science
Scope Classification	Haematology
Services available to the public ¹	No

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered		
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)		
	Name	Address
1	Head Office	City Gate, Mahon, Cork

Scope of Accreditation

Head Office

Blood Transfusion Science

Category: A

Medical pathology field - Test	Test/assay	Specimen Type	Equipment/Technique	Method (CE/Non-CE/In house developed/based on standard method)	Range of measurement	Std. ref & SOP
1020 Transfusion science - .01 Blood grouping including ABO, Rh(D) and other antigens by manual methods	Grouping of blood - ABO, Rh(D) and other antigens by manual methods	Red cells/ Plasma/ Serum	Bio-Rad gel technique Manual tube technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-003
1020 Transfusion science - .02 Blood grouping including ABO, Rh(D) and other antigens by automated methods	Biorad	Whole blood	IH500 automated blood group analyser	CE	N/A	MPC-PP-BT-001
1020 Transfusion science - .03 Blood group antibody screen	Antibody Screening	Plasma	Bio-Rad Manual gel technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-003
1020 Transfusion science - .04 Identification of blood group antibodies	Antibody identification		Bio-Rad Manual gel technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-003
	Biorad	Whole blood	IH500 automated blood group analyser	CE	N/A	MPC-PP-BT-001

1020 Transfusion science - .05 Cross match compatible donor units	Compatibility testing	Donor cells/ plasma	Bio-Rad Manual gel technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-007
1020 Transfusion science - .06 Red cell phenotyping	Red cell phenotyping	Red cells	Bio-Rad Manual gel technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-003
1020 Transfusion science - .09 Direct antiglobulin test	Biorad	Whole blood	IH500 automated blood group analyser	CE	N/A	MPC-PP-BT-001
	Direct Antiglobulin test	Red cells	Bio-Rad Manual gel technique	CE	N/A	ISO 15189:2022 MPC-PP-BT-003
<i>The hospital blood bank has been assessed and is competent to comply with Articles 14 and 15 of the EU Directive 2002/98/EC (S.I. 360/2005 and S.I. 547/2006).</i>						

Head Office

Haematology

Category: A

Medical pathology field - Test	Test/Assay	Specimen Type	Technique	Range of Measurement/Equipment	Method (CE/Non-CE/In house developed/based on standard method)	Std. Ref & SOP
1030 Haematology - .01 Blood counts	Blood count	Whole blood (EDTA)	Flow cytometry Spectrophotometry	WBC (0.00-246 X 10 ³ /μL) RBC (0.00-7.50 X 10 ⁶ /μL) HGB (0.00-25.0 g/dL) HCT (0.00-99.5%) PLT (0.00-3000 X 10 ³ /μL) Cell Dyn Ruby	CE	ISO 15189:2022 MPC-PP-LAB-019
	Full blood count		Flow Cytometry/ Spectrophotometry/ Calculation	Equipment - Cell Dyn Ruby 2	CE	MPC-PP-LAB-019
1030 Haematology - .05 Automated differential leucocyte counts	Automated Differential Leucocyte Counts		Flow cytometry Spectrophotometry	As for WBC Cell Dyn Ruby	CE	ISO 15189:2022 MPC-PP-LAB-019
			Flow Cytometry/ Spectrophotometry	Equipment - Cell Dyn Ruby 2	CE	MPC-PP-LAB-019