

Schedule of Accreditation



Organisation Name	National Virus Reference Laboratory
Trading As	
INAB Reg No	326MT
Contact Name	Eimear Malone
Address	University College Dublin, Belfield, Dublin, D4
Contact Phone No	01-7161319
Email	eimear.malone@ucd.ie
Website	
Accreditation Standard	EN ISO 15189
Standard Version	2022
Date of award of accreditation	09/09/2014
Scope Classification	Microbiology and Virology
Services available to the public ¹	No

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)

	Name	Address
1	Head Office	University College Dublin, Belfield, Dublin, D4

Scope of Accreditation

Head Office

Microbiology and Virology

Category: A

Medical pathology field - Test	Test/assay	Specimen Type	Equipment/Technique	Method (CE/Non-CE/In house developed/based on standard method)	Range of measurement	Std. ref & SOP
1015 Detection and/or identification of bacterial, parasite, fungal and viral nucleic acids - .03 Nucleic acid amplification tests, CE marked commercial systems	Altona RealStar JCV PCR assay	Serum, plasma and CSF	ABI 7500 FAST Analyser	CE Marked	Detected IU/ml and Not Detected	LP/MVIR/152
	Altona SARS-CoV-2 RT-PCR	Respiratory swabs	MagNa Pure 96	CE Marked	Qualitative	LP-CL3-065
	Aptima Mycoplasma Genitalium Assay	Aptima collection device - Anal swab, endocervical swab, urine	Hologic Panther System	CE Marked	Qualitative	VAL/MVIR/225 and LP/MVIR/186

BioFire Filmarray Respiratory Panel	Respiratory Aspirates, BALS, Respiratory Swabs	Biofire Filmarray Instrument	CE Marked Commercial	Qualitative	LP/MVIR/183
BioFire® Respiratory Panel 2.1 plus	Nasopharyngeal Swab	BioFire FilmArray	CE	Detected/Not Detected	LP/MVIR/183
CT/NG	Urine and swabs	Panther	CE	Not Detected RLU < 85 Equivocal 85-250 Detected 250-<4500	LP/MVIR/186
Detection of Adenovirus using the Altona RealStar Adenovirus PCR Kit 1.0	Serum/Plasma	PSH, Magnapure 96, Qiagility, Taqman 7500 FAST	CE Marked Commercial	10*3 - 10*8 copies/ml	LP/MVIR/200
Detection of viruses in CSF samples using the Altona RealStar HSV, VZV & Entero PCR kits	CSF, Serum, Plasma	EasyMAG ® Automated Extractor, Qiagility	CE Marked	Detected, Not Detected	LP/MVIR/220
GeneXpert HIV-1 VL assay	Plasma	GeneXpert Dx system	CE Marked	Detected copies/ml and Not detected	LP/MVIR/173
HCV II Genotyping	Human serum and plasma	Abbott real time assay	CE	> 500 IU/ml Required for Genotyping	LP/MVIR/128
HCV RNA		Abbott real time assay	CE	Quantitative Range 12-10*8 IU/ml	LP/MVIR/125
Hepatitis B Virus		Abbott real time assay	CE	Quantitative Range 10-10*9 IU/ml	LP/MVIR/135
Immunodeficiency virus type 1		Abbott real time assay	CE	Quantitative Range 40-10*7 Copies/ml	LP/MVIR/120
RealStar Altona HHV-6 Assay	Plasma, saliva, swab, CSF	Taqman 7500 Fast	CE Marked	Qualitative	LP/MVIR/112 and VAL/MVIR/233
RealStar Pneumocystis jirovecii PCR kit (Altona)	Sputum, BAL, Aspirate	Taqman 7500	CE	Detected/Not Detected	LP-MVIR-222

Real-Time PCR for BK virus DNA (Quantitative)	Serum, Plasma, and Urine samples	Taqman 7500	CE	Quantitative range 1000- 10*9 Copies/ml	LP/MVIR/177
Real-Time PCR for Hepatitis E Virus	Serum	Taqman 7500	CE marked with validated modification	Positive HEV RNA results > 1x10(4) IU/ml reported quantitatively. b) Samples with a viral load greater than 1x10(10) IU/ml. this reported as > 10.0 log IU/ml. c)Positive HEV RNA results below 1x10(4) IU/ml reported as "DETECTED < 10,000 IU/ml (< 4.0 LogIU/ml)" d) Negative HEV RNA results are reported as "Not detected"	LP/MVIR/160
Real-time PCR for Herpes Simplex 1 and 2 virus DNA	Swabs,	Taqman 7500 analyser and ViiA 7	in-house developed	cT< 35 DETECTED cT 35-38 DETECTED at low level cT > 38, repeat	LP/MVIR/107
Real-time PCR for JC Virus DNA	CSF, Serum, Urine and Plasma	Taqman 7500	CE	Blood / CSF - quantitatively as (a) DETECTED < 2.0 Log10 IU/mL (b) If between 90 and 9e10 IU/mL then report number (c) DETECTED > 11.0 Log10 IU/mL	LP/MVIR/152
		Taqman 7500	CE	Blood / CSF - quantitatively as (a) DETECTED < 2.0	LP/MVIR/152

					Log10 IU/mL (b) If between 90 and 9e10 IU/mL then report number (c) DETECTED > 11.0 Log10 IU/mL	
	Real-Time RT-PCR for Enterovirus and Parechovirus RNA	Stool, Swab	Taqman 7500	in-house developed	Detected cT < 38	0 LP/MVIR/134
	RT PCR for the detection and quantification of HIV type 1 (HIV-1)	Plasma	GeneXpert Instrument	CE Marked Commercial	Quantitative Copies /ml	LP/MVIR/173
	The GeneXpert HCV Viral Load Assay	Plasma	Cepheid GeneXpert IV System	CE Marked	Qualitative range reported as detected <1.08 log IU/ml. Quantitative value reported >1.0 to 8 log IU/ml.	LP-MVIR-173
		Serum and plasma	Cepheid GeneXpert	CE marked with validated modification	HCV genotypes 1-6 over the range of 10 to 100,000,000 IU/mL	LP-MVIR-173
	Trichomonas vaginalis on the Panther	Aptima Collection Device	Panther System	CE Marked Commercial	Qualitative	LP/MVIR/097
1015 Detection and/or identification of bacterial, parasite, fungal and viral nucleic acids - .04 Nucleic acid amplification tests, in house developed assays	In-House HIV Genotypic Antiretroviral Resistance Assay	Plasma	Applied Biosystems 3500 Genetic Analyser	In House	N/A	LP/MVIR/176
	One-Step Real-Time RT-PCR for the Detection of Mumps Virus	Oral Fluid, CSF, other sample types are tested with an NVRL comment	Qiagen, Taqman 7500 FAST	In House	Qualitative	LP/MVIR/112
	Real-Time PCR for Adenovirus DNA	Swabs, Stool	Taqman 7500 analyser and/or Viia 7	in-house developed	cT< 35 DETECTED cT 35-38 DETECTED	LP/MVIR/107

					at low level cT > 38, repeat	
	Real-time PCR for the detection of Norovirus, Astrovirus, Sapovirus, Adenovirus type F, Rotavirus and Rotarix (Gastro screen)	Faecal	Roche MagNA Pure 96, ABI 7500/ViiA7	In House	Detected cT ≤35	LP/MVIR/153
	Real-Time PCR for Varicella Zoster DNA	Swabs,	Taqman 7500 analyser and ViiA 7	in-house developed	cT< 35 DETECTED cT 35-38 DETECTED at low level cT > 38, repeat	LP/MVIR/107
1015 Detection and/or identification of bacterial, parasite, fungal and viral nucleic acids - .05 Nucleotide sequencing & analysis	HBV Genotyping/Resistance	Serum/Plasma	Manual Qiagen, RT-PCR (nested), Sanger sequencing (ABI 3500Dx)	In House	>=log 3 IU/ml	LP/MVIR/193
	HCV Genotyping by Sanger Sequencing		Manual Qiagen, RT-PCR (nested), Sanger sequencing (ABI 3500Dx)	In House	>=log 4 IU/ml	LP/MVIR/197
1018 Detection of antibody response to infection using appropriate CE marked commercial techniques - .01 Particle agglutination, using CE marked commercial systems	TP-PA	Human serum and plasma	Manual	CE	Negative Indeterminate Pos 1:80, 1:160, 1:320, 1:640, 1:1280	LP/VSER/008
1018 Detection of antibody response to infection using appropriate CE marked commercial techniques - .02 Enzyme immunoassay, using CE	Anti-HBc		Vidas	CE	Negative ≥1.4 Equivocal ≥1 to <1.4 Positive <1	LP/VSER/001

marked commercial systems						
	CMV IgM		Vidas	CE	Negative <0.7 Equivocal ≥0.7 to <0.9 Positive ≥0.9	LP/VSER/001
	Cytomegalovirus IgM		Manual	CE	< Cut-off (CO) Negative Lower CO to Upper CO Equivocal ≥Upper CO Positive	LP/VSER/075
	Genscreen Ultra HIV Ag/Ab		Manual	CE	≤0.899 Negative 0.900-1.000 Equivocal 1.001-2.999 Weak Positive ≥3.00 Positive	LP/VSER/075
	HAV IgM		Vidas	CE	Negative <0.4 Equivocal ≥0.4 to <0.499 Positive ≥0.500	LP/VSER/001
	Innotest HCV	Serum/Plasma	ELISA	CE Marked Commercial	Neg (-/=0.800, -/=3.000)	LP/VSER/075
	Leptospira IgM	Human serum	Manual	CE	≤0.899 Negative 0.900-1.100 Equivocal 1.101-1.999 Weak Positive ≥2.000 Positive	LP/VSER/075
	Liaison HEV IgM assay	Serum or Plasma	DiaSorin Liaison® XL	CE	Negative =<0.649, Equivocal 0.650-0.999, Weak Positive 1.000 - 2.999, Positive ≥3.000	LP-VSER-077

Liaison Treponema Screen Assay	Serum and plasma	Diasorin Liaison XL	CE Marked	Qualitative	VAL/VSER/166 and LP/VSER/077
Liaison XL Anti-HBc assay	Serum/Plasma	Chemiluminescence immunoassay (CLIA) technology for the qualitative determination of total antibodies to hepatitis B core antigen (anti-HBc) in human serum or plasma samples	CE	Samples with anti-HBc levels between an index value of ≥ 0.300 -0.900 are graded weak positive. Samples with an index value of 0.300 are graded positive	The manufacturer's direction circular and the NVRL procedure LP/VSER/077 Operation and Maintenance of the Diasorin Liaison XL
Liaison XL Murex Anti-HDV Assay	Serum and plasma	Diasorin Liaison XL	CE Marked	Qualitative	VAL/VSER/162 and LP/VSER/077
Measles IgM	Human serum and plasma	Manual	CE	Serum ≤ 0.899 Negative 0.900-1.099 Equivocal 1.100-1.999 Weak Positive ≥ 2.000 Positive	LP/VSER/075
Monolisa HCV Ab	Human serum and plasma	Manual	CE	≤ 0.899 Negative 0.900-1.000 Equivocal 1.001-2.999 Weak Positive ≥ 3.00 Positive	LP/VSER/075
Mumps IgG		Diasorin Liaison XL	CE	Negative ≤ 9 Equivocal ≥ 9.0 to ≤ 10.99 Weak positive ≥ 11.0 to ≤ 49.99 Positive ≥ 50	LP/VSER/077
Mumps IgM	Human serum and plasma	Manual	CE	Serum	LP/VSER/075

				<0.856 Negative ≥ 0.856 - <1.000 Equivocal ≥ 1.000 - <2.000 Weak Positive ≥2.000 Positive	
Murex HBsAg	Human serum and plasma	Manual	CE	≤0.899 Negative 0.9 – 1.000 Equivocal 1.001-2.999 Weak Positive ≥3.000 Positive	LP/VSER/075
Murex HBsAg Confirmatory		Manual	CE	Positive/Negative	LP/VSER/075
Murex HTLV I/II		Manual	CE	≤0.899 Negative 0.900-1.000 Equivocal 1.001-2.999 Weak Positive ≥3.00 Positive	LP/VSER/075
Ortho HCV		Manual	CE	≤0.899 Negative 0.900-1.000 Equivocal 1.001-2.999 Weak Positive ≥3.00 Positive	LP/VSER/075
Syphilis IgM		Manual	CE	≤0.899 Negative 0.9-1.000 Equivocal 1.001-1.999 Weak Positive ≥2.000 Positive	LP/VSER/075
Toxo IgG II		Vidas	CE	Negative <4 Equivocal ≥4 to <8	LP/VSER/001

					Positive ≥8	
	Toxo IgM		Vidas	CE	Negative <0.55 Equivocal ≥0.55 to <0.65 Positive ≥6.5	LP/VSER/001
	VIDAS Anti-HEV IgM assay	Serum and plasma	BioMerieux VIDAS	CE	Positive Negative	LP-VSER-001
	VIDAS CMV IgG Assay	Serum	BioMerieux VIDAS	CE Marked	Quantitative	LP/VSER/001
1018 Detection of antibody response to infection using appropriate CE marked commercial techniques - .04 Line immunoassay, using CE marked commercial systems	The INNO-LIA Syphilis Score	Serum	Auto Lia 48 Instrument	CE Marked	Positive, Negative, Indeterminate	LP-VSER-016
1018 Detection of antibody response to infection using appropriate CE marked commercial techniques - .07 Chemiluminescent microparticle immunoassay, using CE marked commercial systems	Anti-HBc IgM	Human serum and plasma	Architect i4000	CE	S/CO ≤ 0.999 Negative S/CO 1.000 – 2.999 Weak Positive S/CO ≥ 3.000 Positive	LP/VSER/050
	Anti-HBc II		Architect i4000	CE	S/CO ≤ 0.899 Negative S/CO 0.900 - 1.000 Equivocal S/CO 1.001 - 1.999 Weak Positive S/CO ≥ 2.000 Positive	LP/VSER/050

Anti-HBe
Anti-HBs
Anti-HCV
Borrelia IgG
CMV IgG

Architect i4000	CE	S/CO ≥ 1.101 Negative S/CO 1.002 - 1.100 Equivocal S/CO 0.800 - 1.001 Weak Positive S/CO ≤ 0.800 Positive	LP/VSER/050
Architect i4000	CE	0.0 - 0.49 mIU/ml No antibodies detected 0.50 – 9.99 mIU/ml Less than 10 mIU/ml ≥ 10 mIU/mL mIU/ml value reported	LP/VSER/050
Diasorin Liaison XL	CE	≤ 10 IU/ml 10.1-99.9 mIU/ml ≥ 100 IU/ml	LP/VSER/077
Architect i4000	CE	S/CO ≤ 0.799 Negative S/CO 0.800 - 1.000 Equivocal S/CO 1.001 - 4.999 Weak Positive S/CO ≥ 5.000 Positive	LP/VSER/050
Diasorin Liaison XL	CE	Negative ≤ 9.99 Equivocal ≥ 10.0 to ≤ 14.99 Weak positive ≥ 15.0 to ≤ 29.99 Positive ≥ 30	LP/VSER/077
Architect i4000	CE	≤ 5.999 AU/mL Negative 6.000 – 14.999 AU/mL	LP/VSER/050

CMV IgM
EBNA IgG
EBV IgM
EBV VCA IgG
HAVAb-IgG

		Weak Positive ≥ 15.000 AU/mL Positive	
Architect i4000	CE	Index ≤ 0.849 Negative Index 0.850 – 0.999 Equivocal Index 1.000 – 1.999 Weak Positive Index ≥2.000 Positive	LP/VSER/050
Diasorin Liaison XL	CE	Negative ≤4.99 Equivocal ≥5.00 to ≤20.00 Weak positive ≥20.01 to ≤30.00 Positive ≥30.01	LP/VSER/077
Diasorin Liaison XL	CE	Negative ≤19.9 Equivocal ≥20.0 to ≤39.9 Weak Positive ≥40.0 to ≤79.9 Positive ≥80	LP/VSER/077
Diasorin Liaison XL	CE	Negative ≤19.9 Equivocal ≥20.0 to ≤29.9 Weak positive ≥30.0 to ≤39.9 Positive ≥40	LP/VSER/077
Architect i4000	CE	S/CO ≤ 0.899 Negative S/CO 0.900 - 1.000 Equivocal	LP/VSER/050

HAVAb-IgM
HBeAg
HBsAg
HCV Ag

		S/CO 1.001 - 1.999 Weak Positive S/CO $\geq$ 2.000 Positive	
Architect i4000	CE	S/CO $\leq$ 0.799 Negative S/CO 0.800 - 1.199 Equivocal S/CO 1.200 – 2.999 Weak Positive S/CO $\geq$ 3.000 Positive	LP/VSER/050
Architect i4000	CE	S/CO $\leq$ 0.899 Negative S/CO 0.900 - 1.000 Equivocal S/CO 1.001 - 1.100 Weak Positive S/CO $\geq$ 1.101 Positive	LP/VSER/050
Architect i4000	CE	S/CO $\leq$ 0.799 Negative S/CO 0.800 - 1.000 Equivocal S/CO 1.001 - 14.999 Weak Positive S/CO $\geq$ 15.000 Positive	LP/VSER/050
Architect i4000	CE	0.000 – 2.999 fmol/L Negative 3.000 – 9.999 fmol/L Weak Positive $\geq$ 10.000 fmol/L Positive	LP/VSER/050

HIV Ag/Ab Combo		Architect i4000	CE	S/CO ≤ 0.799 Negative S/CO 0.800 - 1.000 Equivocal S/CO 1.001 – 14.999 Weak Positive S/CO ≥ 15.000 Positive	LP/VSER/050
HSV 1 / 2 IgG	Human serum	Diasorin Liaison XL	CE	Negative ≤ 0.89 Equivocal ≥ 0.9 to ≤ 1.09 Weak positive ≥ 1.10 to ≤ 1.49 Positive ≥ 1.5	LP/VSER/077
Liaison XL Murex recHTLV-I/II assay	Serum/Plasma	Chemiluminescence immunoassay (CLIA) technology for the qualitative determination of specific antibodies to Human T-cell Lymphotropic Virus (HTLV) Type I and Type II (anti-HTLV I and anti-HTLV II) in human serum and plasma samples	CE	Specimens with signal-to-cut-off (S/CO) ratios < 1.00 are considered non-reactive for HTLV I/II antibodies. Specimens with signal-to-cut-off (S/CO) ratios of ≥ 1.00 -2.99 are weak positive and ≥ 3.00 are considered positive for HTLV I/II antibodies	The manufacturer's direction circular and the NVRL procedure LP/VSER/077 Operation and Maintenance of the Diasorin Liaison XL
Measles IgG	Human serum and plasma	Diasorin Liaison XL	CE	Negative < 13.5 Equivocal ≥ 13.5 to ≤ 16.49 Weak Positive ≥ 16.5 to ≤ 49.99 Positive ≥ 50	LP/VSER/077

Parvovirus B19 IgG
Parvovirus B19 IgM
rHTLV-I/II
Rubella IgG
Rubella IgM

Diasorin Liaison XL	CE	Negative <0.9 Equivocal ≥0.9 to <1.1 Weak positive ≥1.1 to <1.5 Positive ≥1.5	LP/VSER/077
Diasorin Liaison XL	CE	Negative <0.9 Equivocal ≥0.9 to <1.1 Weak positive ≥1.1 to <1.5 Positive ≥1.5	LP/VSER/077
Architect i4000	CE	S/CO ≤ 0.799 Negative S/CO 0.800 - 1.000 Equivocal S/CO 1.001 - 1.999 Weak Positive S/CO ≥ 2.000 Positive	LP/VSER/050
Architect i4000	CE	≤ 4.999 IU/ml No antibodies detected 5.000 – 14.999 IU/ml As per algorithm ≥ 15.000 IU/mL Greater than 15 IU/ml	LP/VSER/050
Diasorin Liaison XL	CE	Less than 5 IU/ml <5 IU/ml Detected at low level ≥5.1 to ≤9.999 IU/ml Greater than 10IU/ml ≥10 IU/ml	LP/VSER/077
Architect i4000	CE	S/CO ≤ 0.749 Negative S/CO 0.750 – 1.000	LP/VSER/050

				Equivocal S/CO 1.001 – 1.999 Weak Positive S/CO ≥ 2.000 Positive	
	Syphilis TP		Architect i4000	CE S/CO ≤ 0.499 Negative S/CO 0.500 - 1.000 Equivocal S/CO 1.001 - 3.999 Weak Positive S/CO ≥ 4.000 Positive	LP/VSER/050
	Toxo IgM		Architect i4000	CE S/CO ≤ 0.829 Negative S/CO 0.830 – 0.999 Equivocal S/CO 1.000 – 1.499 Weak Positive S/CO ≥ 1.500 Positive	LP/VSER/050
	ToxoIgG		Architect i4000	CE ≤ 1.599 IU/ml Negative 1.600 – 2.999 IU/mL Equivocal ≥ 3.000 IU/mL Positive	LP/VSER/050