

Schedule of Accreditation



Organisation Name	Airmid Health Laboratory
Trading As	
INAB Reg No	284T
Contact Name	James Coughlan
Address	2056 Castle Drive, Citywest Business Campus, Dublin, D24 YH58
Contact Phone No	016336820
Email	jcoughlan@airmidhealthgroup.com
Website	http://www.airmidhealthgroup.com/
Accreditation Standard	EN ISO/IEC 17025 T
Standard Version	2017
Date of award of accreditation	31/05/2011

¹ Refer to document on interpreting INAB Scopes of Accreditation

Sites from which accredited services are delivered
(the detail of the accredited services delivered at each site are on the Scope of Accreditation)

	Name	Address
1	Head Office	2056 Castle Drive, Citywest Business Campus, Dublin, Ireland, D24 YH58

Scope of Accreditation

Head Office

Chemical Testing

Category: A

Chemistry Field - Tests	Test name	Analyte	Range of measurement	Matrix	Equipment/technique	Standard reference/SOP
797 Miscellaneous materials and products - .03 Other tests	Der p 1 ELISA for quantification of Der p 1 allergen from house dust mites, Dermatophagoides pteronyssinus, in sample extracts (dust, filters and textiles)	Der p 1	(i) 0.625 to 80 ng/ml (ii) 0.016 to 2.0 µg/g (iii) 0.0025 to 0.32 µg Total Recovered Allergen (TRA) (iv) 0.0031 to 0.4 µg TRA (v) 0.0016 to 0.2 µg TRA (vi) 0.156 to 20 µg TRA	(i) Extract (ii) Dust (0.1 g): standard extraction vol. (2.5 ml PBST) (iii) Fibre filter: standard extraction vol. (4 ml PBST) (iv) Mitest filter: (>half full) standard extraction vol. (5 ml PBST) (v) Mitest filter: (<half full) standard extraction vol. (2.5 ml PBST)	ELISA	In-house method SP300 based on Der p 1 ELISA kit (Indoor Biotechnologies)

				(vi) Appliance filter (10x15 cm): standard extraction vol. (250 ml PBST)		
	<p>Fel d 1 ELISA for quantification of Fel d 1 allergen from domestic cats, <i>Felis domesticus</i>, in sample extracts (dust, filters and textiles)</p>	<p>Fel d 1</p>	<p>(i) 0.078 to 20 ng/ml (ii) 0.002 to 0.5 µg/g (iii) 0.0003 to 0.08 µg TRA (iv) 0.0004 to 0.1 µg TRA (v) 0.0002 to 0.05 µg TRA (vi) 0.195 to 5 µg TRA</p>	<p>(i) Extract (ii) Dust (0.1 g): standard extraction vol. (2.5 ml PBST) (iii) Fibre filter: standard extraction vol. (4 ml PBST) (iv) Mitest filter: (>half full) standard extraction vol. (5 ml PBST) (v) Mitest filter: (<half full) standard extraction vol. (2.5 ml PBST) (vi) Appliance filter (10x15 cm): standard extraction vol. (250 ml PBST)</p>	<p>ELISA</p>	<p>In-house method SP301 based on Fel d 1 ELISA kit (Indoor Biotechnologies)</p>
	<p>qPCR Method for quantification of Influenza A (H1N1) virus in sample extracts.</p>	<p>Influenza A (H1N1)</p>	<p>2 - 200000 copies per µL</p>	<p>Extract</p>	<p>qPCR</p>	<p>In-house method SP622 based on Human Influenza A virus (M1 gene) Genesig Advanced Kit by Primerdesign.</p>

